

LB 112873


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

36215



A= 0.0000e+000

B= 2.7708e-004

C= -2.0491e-002

Rho= 0.9998829

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.005	-0.005	57	0.000	57					-0
0.05	0.050	0.050	0.000	255	0.0 %	255	0				-9
0.2	0.200	0.183	-0.017	735	0.0 %	735					-1
2.5	2.500	2.485	-0.015	9044	0.0 %	9044					-1
5.0	5.000	5.115	0.115	18534	0.0 %	18534					2
7.5	7.500	7.407	-0.093	26806	0.0 %	26806					-1
10.0	10.000	10.014	0.014	36215	0.0 %	36215					0

LB112873 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 21 Jan 2021 13:23:55

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	57	-	0.0000	245.1	PPB	21 Jan 2021 13:29:53	S	Std
0.05 - 1	50.05	255	-	0.0500	245.1	PPB	21 Jan 2021 13:32:10	S	Std
0.2 - 1	50.2	735	-	0.2000	245.1	PPB	21 Jan 2021 13:34:27	S	Std
2.5 - 1	52.5	9044	-	2.5000	245.1	PPB	21 Jan 2021 13:36:44	S	Std
5.0 - 1	55	18534	-	5.0000	245.1	PPB	21 Jan 2021 13:39:01	S	Std
7.5 - 1	57.5	26806	-	7.5000	245.1	PPB	21 Jan 2021 13:41:16	S	Std
10.0 - 1	510	36215	-	10.0000	245.1	PPB	21 Jan 2021 13:46:10	S	Std
ICV78 - 1	ICV78	13928	3.8387	-245.1		PPB	21 Jan 2021 13:49:00	U	SMPL
ICB78 - 1	ICB78	-48	-0.0338	-245.1		PPB	21 Jan 2021 13:51:15	U	SMPL
CCV58 - 1	CCV58	18442	5.0895	-245.1		PPB	21 Jan 2021 13:53:29	U	SMPL
CCB58 - 1	CCB58	-60	-0.0371	-245.1		PPB	21 Jan 2021 13:55:45	U	SMPL
CRA - 1	CRA	234	0.0443	-245.1		PPB	21 Jan 2021 13:58:00	U	SMPL
HighStd - 1	HighStd	36486	10.0891	-245.1		PPB	21 Jan 2021 14:00:15	U	SMPL
ChkStd - 1	ChkStd	26003	7.1845	-245.1		PPB	21 Jan 2021 14:02:30	U	SMPL
PB134259BL - 1	PBW	-99	-0.0479	-245.1		PPB	21 Jan 2021 14:04:48	U	SMPL
PB134259BS - 1	LCSW	12810	3.5289	-245.1		PPB	21 Jan 2021 14:07:04	U	SMPL
M1101-01 - 1	55-ST-30	-4	-0.0216	-245.1		PPB	21 Jan 2021 14:09:20	U	SMPL
M1126-05 - 1	PT-HG-WS	30136	8.3296	-245.1		PPB	21 Jan 2021 14:11:37	U	SMPL
M1175-01 - 1	MANHOLE	193	0.0330	-245.1		PPB	21 Jan 2021 14:13:54	U	SMPL
M1175-01DUP - 1	MANHOLEDUP	239	0.0457	-245.1		PPB	21 Jan 2021 14:16:12	U	SMPL
M1175-01MS - 1	MANHOLEMS	12239	3.3707	-245.1		PPB	21 Jan 2021 14:18:27	U	SMPL
CCV59 - 1	CCV59	18864	5.2064	-245.1		PPB	21 Jan 2021 14:20:43	U	SMPL
CCB59 - 1	CCB59	-69	-0.0396	-245.1		PPB	21 Jan 2021 14:22:59	U	SMPL
M1175-01MSD - 1	MANHOLEMSD	12941	3.5652	-245.1		PPB	21 Jan 2021 14:25:15	U	SMPL
M1175-01LX5 - 1		-34	-0.0299	-245.1		PPB	21 Jan 2021 14:39:17	U	SMPL
M1175-01A - 1		14475	3.9903	-245.1		PPB	21 Jan 2021 14:41:33	U	SMPL
CCV60 - 1	CCV60	18688	5.1576	-245.1		PPB	21 Jan 2021 14:43:50	U	SMPL
CCB60 - 1	CCB60	-56	-0.0360	-245.1		PPB	21 Jan 2021 14:46:07	U	SMPL