



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112876

Slope : 98.9

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	01/21/2021	12:50
2	CAL2	1	Water	NA	NA	21.1	7.00	01/21/2021	12:51
3	CAL3	1	Water	NA	NA	21.1	10.03	01/21/2021	12:53
4	ICV	1	Water	NA	NA	21.2	6.98	01/21/2021	13:00
5	CCV1	1	Water	NA	NA	21.1	2.01	01/21/2021	13:02
6	M1194-01	1	Water	NA	NA	20.1	5.26	01/21/2021	13:05
7	M1194-01DUP	1	Water	NA	NA	20.2	5.27	01/21/2021	13:06
8	M1195-01	1	Water	NA	NA	20.3	4.74	01/21/2021	13:10
9	M1195-02	1	Water	NA	NA	20.1	4.60	01/21/2021	13:12
10	M1195-03	1	Water	NA	NA	20.2	4.86	01/21/2021	13:15
11	M1196-01	1	Water	NA	NA	20.2	4.93	01/21/2021	13:19
12	M1196-02	1	Water	NA	NA	20.1	4.86	01/21/2021	13:20
13	CCV2	1	Water	NA	NA	21.2	12.02	01/21/2021	13:22

2B 12876

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M1194 WorkList ID : 145879 Department : Wet-Chemistry Date : 01-21-2021 12:44:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/20/2021	Water	M1194-01	pH	Cool 4 deg C	STER01	--Select--	POOL-WATER	01/20/2021	9040C
01/20/2021	Water	M1195-01	pH	Cool 4 deg C	STER09	G11	CELL-A-POOL-WATER(DE-IC)	01/20/2021	9040C
01/20/2021	Water	M1195-02	pH	Cool 4 deg C	STER09	G11	CELL-B-C-POOL-WATER-(DE-IC)	01/20/2021	9040C
01/20/2021	Water	M1195-03	pH	Cool 4 deg C	STER09	G11	CELL-D-POOL-WATER-(DE-IC)	01/20/2021	9040C
01/20/2021	Water	M1196-01	pH	Cool 4 deg C	STER16	H21	MAIN-CELL	01/20/2021	9040C
01/19/2021	Water	M1196-02	pH	Cool 4 deg C	STER16	H21	EX-CELL	01/19/2021	9040C

Date/Time 01-21-2021 12:45
 Received by: J. West (202007)
 Relinquished by: J. West (202007)

Date/Time 01-21-2021 14:30
 Received by: J. West (202007)
 Relinquished by: J. West (202007)