



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 1/22/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 15:40
In Date: 01/21/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:10
Out Date: 01/22/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB112877

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M1203-07	X2Y83	1	1.00	2.00	2.00	100.0	P.E.SAMPLE
M1203-08	X2Y84	2	1.00	2.00	2.00	100.0	P.E.SAMPLE
M1203-09	X2Y85	3	1.00	2.00	2.00	100.0	P.E.SAMPLE
M1203-10	X2Y86	4	1.00	2.00	2.00	100.0	P.E.SAMPLE
M1203-11	X2Y87	5	1.00	2.00	2.00	100.0	P.E.SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

VB 112877

WorkList Name : %1-M1203 WorkList ID : 145882 Department : Wet-Chemistry Date : 01-21-2021 15:30:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/30/2021	Solid	M1203-07	Percent Solids	Cool 4 deg C	USEP04	QA Office	X2Y83	01/20/2021	Chemtech -SO
01/30/2021	Solid	M1203-08	Percent Solids	Cool 4 deg C	USEP04	QA Office	X2Y84	01/20/2021	Chemtech -SO
01/30/2021	Solid	M1203-09	Percent Solids	Cool 4 deg C	USEP04	QA Office	X2Y85	01/20/2021	Chemtech -SO
01/30/2021	Solid	M1203-10	Percent Solids	Cool 4 deg C	USEP04	QA Office	X2Y86	01/20/2021	Chemtech -SO
01/30/2021	Solid	M1203-11	Percent Solids	Cool 4 deg C	USEP04	QA Office	X2Y87	01/20/2021	Chemtech -SO

Date/Time 01-21-2021
 Received by: NA
 Relinquished by: NA

Date/Time 01-21-2021
 Received by: NA
 Relinquished by: NA