



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112884

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	01/22/2021	10:25
2	CAL2	1	Water	NA	NA	21.2	7.00	01/22/2021	10:26
3	CAL3	1	Water	NA	NA	21.2	10.03	01/22/2021	10:27
4	ICV	1	Water	NA	NA	21.2	7.02	01/22/2021	10:29
5	CCV1	1	Water	NA	NA	21.2	2.01	01/22/2021	10:30
6	M1189-04	1	Solid	20.03	20	20.2	9.06	01/22/2021	10:33
7	M1189-05	1	Solid	20.04	20	20.1	9.06	01/22/2021	10:34
8	M1209-01	1	Solid	20.01	20	20.5	6.49	01/22/2021	10:36
9	M1209-01DUP	1	Solid	20.02	20	20.6	6.50	01/22/2021	10:37
10	CCV2	1	Water	NA	NA	21.2	12.02	01/22/2021	10:39

1B 112884

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S M1107 WorkList ID : 145906 Department : Wet-Chemistry Date : 01-22-2021 10:21:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/21/2021	Solid	M1189-04	pH	Cool 4 deg C	PSEG03	H22	41047	01/21/2021	9045D
01/21/2021	Solid	M1189-05	pH	Cool 4 deg C	PSEG03	H22	41048	01/21/2021	9045D
01/21/2021	Solid	M1209-01	pH	Cool 4 deg C	TULL02	G12	FK-PS-01-069	01/21/2021	9045D

Date/Time 01.22.2021 10:25
Received by: ID/est Dean
Relinquished by: ID/est Dean

Date/Time 01.22.2021 11:35
Received by: ID/est Dean
Relinquished by: ID/est Dean