



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112885

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	01/22/2021	10:25
2	CAL2	1	Water	NA	NA	21.2	7.00	01/22/2021	10:26
3	CAL3	1	Water	NA	NA	21.2	10.03	01/22/2021	10:27
4	ICV	1	Water	NA	NA	21.2	7.02	01/22/2021	10:29
5	CCV1	1	Water	NA	NA	21.2	2.01	01/22/2021	10:30
6	M1189-06	1	Water	NA	NA	20.2	5.96	01/22/2021	10:40
7	M1189-08	1	Water	NA	NA	20.1	5.93	01/22/2021	10:42
8	M1189-09	1	Water	NA	NA	20.2	6.64	01/22/2021	10:44
9	M1189-13	1	Water	NA	NA	20.1	6.43	01/22/2021	10:45
10	M1189-14	1	Water	NA	NA	20.1	6.50	01/22/2021	10:48
11	M1189-15	1	Water	NA	NA	20.2	6.20	01/22/2021	10:50
12	M1189-16	1	Water	NA	NA	20.2	6.36	01/22/2021	11:00
13	M1189-17	1	Water	NA	NA	20.1	6.92	01/22/2021	11:02
14	M1189-18	1	Water	NA	NA	20.2	8.56	01/22/2021	11:05
15	M1189-19	1	Water	NA	NA	20.7	8.22	01/22/2021	11:06
16	CCV2	1	Water	NA	NA	21.3	12.02	01/22/2021	11:11
17	M1189-20	1	Water	NA	NA	20.3	7.36	01/22/2021	11:12
18	M1192-04	1	Water	NA	NA	20.1	6.43	01/22/2021	11:15
19	M1192-05	1	Water	NA	NA	20.3	6.81	01/22/2021	11:17
20	M1192-05DUP	1	Water	NA	NA	20.4	6.82	01/22/2021	11:18

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water	NA	NA	21.2	2.01	01/22/2021	11:20

2-B 112885

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M1192 WorkList ID : 145907 Department : Wet-Chemistry Date : 01-22-2021 10:22:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/21/2021	Water	M1189-06	pH	Cool 4 deg C	PSEG03	H22	41041	01/21/2021	9040C
01/21/2021	Water	M1189-08	pH	Cool 4 deg C	PSEG03	H22	41051	01/21/2021	9040C
01/21/2021	Water	M1189-09	pH	Cool 4 deg C	PSEG03	H22	41052	01/21/2021	9040C
01/21/2021	Water	M1189-13	pH	Cool 4 deg C	PSEG03	H22	41089	01/21/2021	9040C
01/21/2021	Water	M1189-14	pH	Cool 4 deg C	PSEG03	H22	41090	01/21/2021	9040C
01/21/2021	Water	M1189-15	pH	Cool 4 deg C	PSEG03	H22	41091	01/21/2021	9040C
01/21/2021	Water	M1189-16	pH	Cool 4 deg C	PSEG03	H22	41105	01/21/2021	9040C
01/21/2021	Water	M1189-17	pH	Cool 4 deg C	PSEG03	H22	41106	01/21/2021	9040C
01/21/2021	Water	M1189-18	pH	Cool 4 deg C	PSEG03	H22	41107	01/21/2021	9040C
01/21/2021	Water	M1189-19	pH	Cool 4 deg C	PSEG03	H22	411087	01/21/2021	9040C
01/21/2021	Water	M1189-20	pH	Cool 4 deg C	PSEG03	H22	41088	01/21/2021	9040C
01/21/2021	Water	M1192-04	pH	Cool 4 deg C	PSEG03	H21	41623	01/21/2021	9040C
01/21/2021	Water	M1192-05	pH	Cool 4 deg C	PSEG03	H21	41046	01/21/2021	9040C

Date/Time 01-22-2021 10:25

Received by: SP [Signature]

Relinquished by: SP [Signature]

Date/Time 01-22-2021 11:35

Received by: SP [Signature]

Relinquished by: SP [Signature]