



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB112887

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	01/22/2021	11:37
2	CAL2	1	Water	NA	NA	21.3	7.00	01/22/2021	11:38
3	CAL3	1	Water	NA	NA	21.3	10.03	01/22/2021	11:39
4	ICV	1	Water	NA	NA	21.3	7.02	01/22/2021	11:40
5	CCV1	1	Water	NA	NA	21.3	2.01	01/22/2021	11:42
6	M1219-01	1	Water	NA	NA	20.2	7.13	01/22/2021	11:45
7	M1219-01DUP	1	Water	NA	NA	20.2	7.14	01/22/2021	11:46
8	M1219-02	1	Water	NA	NA	20.2	7.23	01/22/2021	11:48
9	CCV2	1	Water	NA	NA	21.4	12.02	01/22/2021	11:50

WORKLIST(Hardcopy Internal Chain)

LB 112887

WorkList Name : CORROSIONITY M1219 WorkList ID : 145913 Department : Wet-Chemistry Date : 01-22-2021 11:28:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/21/2021	Water	M1219-01	F Corrosivity	Cool 4 deg C	JACO05	G21	IDW-BB2-AQ-012121-01	01/21/2021	9040C
01/21/2021	Water	M1219-02	F Corrosivity	Cool 4 deg C	JACO05	G21	IDW-BB2-AQ-012121-02	01/21/2021	9040C

Date/Time 01-22-2021 11:30
 Received by: JP West (Dem)
 Relinquished by: JP West (Dem)

Date/Time 01-22-2021 12:40
 Received by: JP West
 Relinquished by: JP West (Dem)