

LB112892


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

50407



A= 0.0000e+000

B= 1.9791e-004 slope

C= -2.0362e-002 y-intercept

Rho= 0.9999594

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.012	0.012	164	0.000	164					=
0.05	0.050					226					=
0.20	0.200	0.187	-0.013	1049	0.0 %	1049					-7
2.50	2.500	2.473	-0.027	12600	0.0 %	12600					-1
5.00	5.000	5.013	0.013	25431	0.0 %	25431					0
7.50	7.500	7.559	0.059	38297	0.0 %	38297					1
10.0	10.000	9.956	-0.044	50407	0.0 %	50407					0

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	164	-		0 ISM02.4	PPB	1/22/2021 13:57	Std
	0.2 S0.2	1049	-		0.2 ISM02.4	PPB	1/22/2021 13:59	Std
	2.5 S2.5	12600	-		2.5 ISM02.4	PPB	1/22/2021 14:02	Std
	5 S5.0	25431	-		5 ISM02.4	PPB	1/22/2021 14:04	Std
	7.5 S7.5	38297	-		7.5 ISM02.4	PPB	1/22/2021 14:08	Std
	10 S10.0	50407	-		10 ISM02.4	PPB	1/22/2021 14:12	Std
ICV	ICV	19646	3.8678 -		ISM02.4	PPB	1/22/2021 14:15	SMPL
ICB	ICB	146	0.0085 -		ISM02.4	PPB	1/22/2021 14:17	SMPL
CCV065	CCV065	25296	4.986 -		ISM02.4	PPB	1/22/2021 14:20	SMPL
CCB065	CCB065	139	0.0071 -		ISM02.4	PPB	1/22/2021 14:22	SMPL
PB134282BL	PBS282	49	-0.0107 -		ISM02.4	PPB	1/22/2021 14:24	SMPL
M1165-01	MF1C00	4964	0.9621 -		ISM02.4	PPB	1/22/2021 14:26	SMPL
M1165-02	MF1C01	8818	1.7248 -		ISM02.4	PPB	1/22/2021 14:29	SMPL
M1165-03	MF1C02	53445	10.5569 -		ISM02.4	PPB	1/22/2021 14:31	SMPL
M1165-04	MF1C03	4096	0.7903 -		ISM02.4	PPB	1/22/2021 14:33	SMPL
M1165-05	MF1C04	5056	0.9803 -		ISM02.4	PPB	1/22/2021 14:35	SMPL
M1165-06	MF1C05	13863	2.7233 -		ISM02.4	PPB	1/22/2021 14:38	SMPL
M1165-07	MF1C06	6873	1.3399 -		ISM02.4	PPB	1/22/2021 14:40	SMPL
M1165-08	MF1C07	29456	5.8093 -		ISM02.4	PPB	1/22/2021 14:42	SMPL
M1165-09	MF1C08	132925	26.2868 -		ISM02.4	PPB	1/22/2021 14:45	SMPL
M1165-10	MF1C09	32948	6.5004 -		ISM02.4	PPB	1/22/2021 14:47	SMPL
M1165-11	MF1C10	39608	7.8185 -		ISM02.4	PPB	1/22/2021 14:49	SMPL
M1165-12	MF1C11	40349	7.9651 -		ISM02.4	PPB	1/22/2021 14:51	SMPL
M1165-13	MF1C12	374	0.0537 -		ISM02.4	PPB	1/22/2021 14:54	SMPL
M1165-14	MF1C13	38912	7.6807 -		ISM02.4	PPB	1/22/2021 14:56	SMPL
M1165-15	MF1C14	32485	6.4087 -		ISM02.4	PPB	1/22/2021 14:58	SMPL
M1165-16	MF1C15	84566	16.7161 -		ISM02.4	PPB	1/22/2021 15:00	SMPL
M1165-17	MF1C16	34416	6.7909 -		ISM02.4	PPB	1/22/2021 15:03	SMPL
M1165-18	MF1C16D	31124	6.1394 -		ISM02.4	PPB	1/22/2021 15:05	SMPL
M1165-19	MF1C16S	46323	9.1474 -		ISM02.4	PPB	1/22/2021 15:10	SMPL
CCV066	CCV066	25645	5.055 -		ISM02.4	PPB	1/22/2021 15:13	SMPL
CCB066	CCB066	98	-0.001 -		ISM02.4	PPB	1/22/2021 15:15	SMPL
M1165-03DLX2	MF1C02	26580	5.2401 -		ISM02.4	PPB	1/22/2021 15:17	SMPL
M1165-09DLX5	MF1C08	26715	5.2668 -		ISM02.4	PPB	1/22/2021 15:19	SMPL
M1165-16DLX5	MF1C15	16516	3.2483 -		ISM02.4	PPB	1/22/2021 15:22	SMPL
CCV067	CCV067	25446	5.0157 -		ISM02.4	PPB	1/22/2021 15:24	SMPL
CCB067	CCB067	126	0.0046 -		ISM02.4	PPB	1/22/2021 15:26	SMPL