



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB112896

Slope : 97.9

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/22/2021	11:00
2	CAL2	1	Water	NA	NA	20.2	7.00	01/22/2021	11:02
3	CAL3	1	Water	NA	NA	20.3	10.02	01/22/2021	11:03
4	ICV	1	Water	NA	NA	20.3	6.98	01/22/2021	11:05
5	CCV1	1	Water	NA	NA	20.2	2.01	01/22/2021	11:06
6	M1188-02	1	Solid	20.03	20	20.4	7.71	01/22/2021	11:10
7	M1188-05	1	Solid	20.01	20	20.2	7.17	01/22/2021	11:12
8	M1213-01	1	Solid	20.03	20	20.2	6.67	01/22/2021	11:15
9	M1213-01DUP	1	Solid	20.04	20	20.4	6.68	01/22/2021	11:16
10	CCV2	1	Water	NA	NA	20.2	12.02	01/22/2021	11:19

WORKLIST(Hardcopy Internal Chain)

VB 112896

WorkList Name : CORROSIONITY M1188 WorkList ID : 145910 Department : Wet-Chemistry Date : 01-22-2021 10:49:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/21/2021	Solid	M1188-02	Corrosivity	Cool 4 deg C	PSEG03	H22	TP-230	01/21/2021	9045D
01/21/2021	Solid	M1188-05	Corrosivity	Cool 4 deg C	PSEG03	H22	TP-SWITCH-GEAR	01/21/2021	9045D
01/21/2021	Solid	M1213-01	Corrosivity	Cool 4 deg C	JACO05	G12	IDW-BR2-SO-012121	01/21/2021	9045D

Date/Time 01-22-2021 10:50
 Received by: JG West
 Relinquished by: JG West

Date/Time 01-22-2021 12:30
 Received by: JG West
 Relinquished by: JG West