

LB112897


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

44235



A= 0.0000e+000

B= 2.3113e-004

C= 1.4059e-002

Rho= 0.9987922

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.038	0.038	102	0.000	102					5
0.2	0.200	0.210	0.010	846	0.0 %	846					4
2.5	2.500	2.605	0.105	11209	0.0 %	11209					1
5.0	5.000	4.966	-0.034	21427	0.0 %	21427					5
7.5	7.500	7.144	-0.356	30847	0.0 %	30847					2
10.0	10.000	10.238	0.238	44235	0.0 %	44235					

LB112897 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Jan 2021 16:03:55

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	102	-	0.0000	7470A	PPB	22 Jan 2021 16:12:14	S	Std
0.2 - 1	50.2	846	-	0.2000	7470A	PPB	22 Jan 2021 16:17:06	S	Std
2.5 - 1	52.5	11209	-	2.5000	7470A	PPB	22 Jan 2021 16:19:23	S	Std
5.0 - 1	55	21427	-	5.0000	7470A	PPB	22 Jan 2021 16:21:40	S	Std
7.5 - 1	57.5	30847	-	7.5000	7470A	PPB	22 Jan 2021 16:23:56	S	Std
10.0 - 1	510	44235	-	10.0000	7470A	PPB	22 Jan 2021 16:31:26	S	Std
ICV80 - 1	ICV80	15996	3.7112	-	7470A	PPB	22 Jan 2021 16:34:20	U	SMPL
ICB80 - 1	ICB80	22	0.0191	-	7470A	PPB	22 Jan 2021 16:39:06	U	SMPL
CCV68 - 1	CCV68	20954	4.8571	-	7470A	PPB	22 Jan 2021 16:41:21	U	SMPL
CCB68 - 1	CCB68	-41	0.0046	-	7470A	PPB	22 Jan 2021 16:43:36	U	SMPL
CRA - 1	CRA	847	0.2098	-	7470A	PPB	22 Jan 2021 16:45:51	U	SMPL
HighStd - 1	HighStd	44565	10.3142	-	7470A	PPB	22 Jan 2021 16:48:06	U	SMPL
ChkStd - 1	ChkStd	30306	7.0186	-	7470A	PPB	22 Jan 2021 16:50:21	U	SMPL
PB134293BL - 1	PBW	-75	-0.0033	-	7470A	PPB	22 Jan 2021 16:52:38	U	SMPL
PB134293BS - 1	LCSW	16005	3.7132	-	7470A	PPB	22 Jan 2021 16:57:30	U	SMPL
M1169-02 - 1	TP-C	178	0.0552	-	7470A	PPB	22 Jan 2021 17:02:20	U	SMPL
M1169-02DUP - 1	TP-CDUP	113	0.0402	-	7470A	PPB	22 Jan 2021 17:04:36	U	SMPL
M1169-02MS - 1	TP-CMS	14487	3.3624	-	7470A	PPB	22 Jan 2021 17:06:53	U	SMPL
M1169-02MSD - 1	TP-CMSD	14558	3.3788	-	7470A	PPB	22 Jan 2021 17:09:11	U	SMPL
CCV69 - 1	CCV69	20963	4.8592	-	7470A	PPB	22 Jan 2021 17:11:26	U	SMPL
CCB69 - 1	CCB69	-49	0.0027	-	7470A	PPB	22 Jan 2021 17:13:42	U	SMPL
PB134293TB - 1	PB134212TB	0	0.0141	-	7470A	PPB	22 Jan 2021 17:15:57	U	SMPL
M1169-02LX5 - 1		19	0.0185	-	7470A	PPB	22 Jan 2021 17:18:13	U	SMPL
M1169-02A - 1		15766	3.6580	-	7470A	PPB	22 Jan 2021 17:20:28	U	SMPL
CCV70 - 1	CCV70	21104	4.8918	-	7470A	PPB	22 Jan 2021 17:22:44	U	SMPL
CCB70 - 1	CCB70	-41	0.0046	-	7470A	PPB	22 Jan 2021 17:25:00	U	SMPL