



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112901

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

True Value of CCV4 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	01/22/2021	17:10
2	CAL2	1	Water	NA	NA	20.1	7.00	01/22/2021	17:11
3	CAL3	1	Water	NA	NA	20.1	10.03	01/22/2021	17:12
4	ICV	1	Water	NA	NA	20.2	7.01	01/22/2021	17:15
5	CCV1	1	Water	NA	NA	20.1	2.01	01/22/2021	17:16
6	M1216-04	1	Water	NA	NA	20.2	12.01	01/22/2021	17:19
7	M1216-04DUP	1	Water	NA	NA	20.3	12.02	01/22/2021	17:20
8	M1217-02	1	Water	NA	NA	20.3	7.51	01/22/2021	17:22
9	M1217-03	1	Water	NA	NA	20.1	7.12	01/22/2021	17:25
10	M1217-04	1	Water	NA	NA	20.5	13.10	01/22/2021	17:30
11	M1217-05	1	Water	NA	NA	20.4	13.48	01/22/2021	17:33
12	M1217-06	1	Water	NA	NA	20.2	6.91	01/22/2021	17:35
13	M1217-07	1	Water	NA	NA	20.2	6.42	01/22/2021	17:36
14	M1217-08	1	Water	NA	NA	20.2	6.21	01/22/2021	17:39
15	M1217-09	1	Water	NA	NA	20.3	6.18	01/22/2021	17:42
16	CCV2	1	Water	NA	NA	20.2	12.02	01/22/2021	17:45
17	M1217-10	1	Water	NA	NA	20.1	6.44	01/22/2021	17:48
18	M1217-11	1	Water	NA	NA	20.3	11.44	01/22/2021	17:50
19	M1217-12	1	Water	NA	NA	20.1	11.69	01/22/2021	17:52
20	M1217-13	1	Water	NA	NA	20.1	7.13	01/22/2021	18:02

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	M1217-14	1	Water	NA	NA	13.0	6.52	01/22/2021	18:05
22	M1217-15	1	Water	NA	NA	20.4	8.20	01/22/2021	18:06
23	M1217-16	1	Water	NA	NA	20.1	9.35	01/22/2021	18:10
24	M1217-17	1	Water	NA	NA	20.1	6.90	01/22/2021	18:11
25	M1217-18	1	Water	NA	NA	20.4	13.18	01/22/2021	18:13
26	M1217-19	1	Water	NA	NA	20.1	7.25	01/22/2021	18:15
27	CCV3	1	Water	NA	NA	20.3	2.01	01/22/2021	18:16
28	M1218-04	1	Water	NA	NA	18.9	6.75	01/22/2021	18:20
29	M1218-06	1	Water	NA	NA	20.1	6.15	01/22/2021	18:21
30	M1218-06DUP	1	Water	NA	NA	20.2	6.16	01/22/2021	18:22
31	CCV4	1	Water	NA	NA	20.4	12.02	01/22/2021	18:25

LB 112901

WORKLIST(Hardcopy Internal Chain)

Date : 01-22-2021 16:35:06

Department : Wet-Chemistry

WorkList ID : 145922

WorkList Name : PH M1206

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/22/2021	Water	M1216-04	pH	Cool 4 deg C	PSEG03	H31	PS-103	01/22/2021	9040C
01/22/2021	Water	M1217-02	pH	Cool 4 deg C	PSEG03	H33	1723	01/22/2021	9040C
01/22/2021	Water	M1217-03	pH	Cool 4 deg C	PSEG03	H33	1724	01/22/2021	9040C
01/22/2021	Water	M1217-04	pH	Cool 4 deg C	PSEG03	H33	1726	01/22/2021	9040C
01/22/2021	Water	M1217-05	pH	Cool 4 deg C	PSEG03	H33	1727	01/22/2021	9040C
01/22/2021	Water	M1217-06	pH	Cool 4 deg C	PSEG03	H33	1674	01/22/2021	9040C
01/22/2021	Water	M1217-07	pH	Cool 4 deg C	PSEG03	H33	1675	01/22/2021	9040C
01/22/2021	Water	M1217-08	pH	Cool 4 deg C	PSEG03	H33	1676	01/22/2021	9040C
01/22/2021	Water	M1217-09	pH	Cool 4 deg C	PSEG03	H33	1677	01/22/2021	9040C
01/22/2021	Water	M1217-10	pH	Cool 4 deg C	PSEG03	H33	1678	01/22/2021	9040C
01/22/2021	Water	M1217-11	pH	Cool 4 deg C	PSEG03	H33	1679	01/22/2021	9040C
01/22/2021	Water	M1217-12	pH	Cool 4 deg C	PSEG03	H33	1680	01/22/2021	9040C
01/22/2021	Water	M1217-13	pH	Cool 4 deg C	PSEG03	H33	1682	01/22/2021	9040C
01/22/2021	Water	M1217-14	pH	Cool 4 deg C	PSEG03	H33	1685	01/22/2021	9040C
01/22/2021	Water	M1217-15	pH	Cool 4 deg C	PSEG03	H33	1525	01/22/2021	9040C
01/22/2021	Water	M1217-16	pH	Cool 4 deg C	PSEG03	H33	1710	01/22/2021	9040C
01/22/2021	Water	M1217-17	pH	Cool 4 deg C	PSEG03	H33	1723-1724	01/22/2021	9040C
01/22/2021	Water	M1217-18	pH	Cool 4 deg C	PSEG03	H33	1726-1727	01/22/2021	9040C
01/22/2021	Water	M1217-19	pH	Cool 4 deg C	PSEG03	H33	1674-1685	01/22/2021	9040C
01/22/2021	Water	M1218-04	pH	Cool 4 deg C	PSEG03	H22	238	01/22/2021	9040C
01/22/2021	Water	M1218-06	pH	Cool 4 deg C	PSEG03	H22	237	01/22/2021	9040C

Date/Time 01.22.2021 16:40

Received by: JP (not copy)

Relinquished by: JP (not copy)

Date/Time

Received by: JP (not copy)

Relinquished by: JP (not copy)

01.22.2021 18:20