



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112910

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/25/2021	15:00
2	CAL2	1	Water	NA	NA	20.2	7.01	01/25/2021	15:02
3	CAL3	1	Water	NA	NA	20.2	10.03	01/25/2021	15:03
4	ICV	1	Water	NA	NA	20.2	6.99	01/25/2021	15:05
5	CCV1	1	Water	NA	NA	20.2	2.01	01/25/2021	15:06
6	M1190-03	1	Water	NA	NA	20.2	6.49	01/25/2021	15:10
7	M1190-04	1	Water	NA	NA	20.1	6.19	01/25/2021	15:13
8	M1190-05	1	Water	NA	NA	20.1	6.16	01/25/2021	15:15
9	M1190-06	1	Water	NA	NA	20.2	6.96	01/25/2021	15:19
10	M1190-07	1	Water	NA	NA	20.3	6.08	01/25/2021	15:22
11	M1190-08	1	Water	NA	NA	20.1	6.80	01/25/2021	15:25
12	M1229-04	1	Water	NA	NA	20.4	6.53	01/25/2021	15:30
13	M1232-01	1	Water	NA	NA	20.1	5.01	01/25/2021	15:32
14	M1232-01DUP	1	Water	NA	NA	20.2	5.02	01/25/2021	15:33
15	CCV2	1	Water	NA	NA	20.3	12.02	01/25/2021	15:35

WORKLIST(Hardcopy Internal Chain)

2B 112910

WorkList Name : ph m1232 WorkList ID : 145947 Department : Wet-Chemistry Date : 01-25-2021 13:37:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/25/2021	Water	M1190-03	✓ pH	Cool 4 deg C	PSEG03	H41	41109	01/25/2021	9040C
01/25/2021	Water	M1190-04	✓ pH	Cool 4 deg C	PSEG03	H41	41054	01/25/2021	9040C
01/25/2021	Water	M1190-05	✓ pH	Cool 4 deg C	PSEG03	H41	41053	01/25/2021	9040C
01/25/2021	Water	M1190-06	✓ pH	Cool 4 deg C	PSEG03	H41	41111	01/25/2021	9040C
01/25/2021	Water	M1190-07	✓ pH	Cool 4 deg C	PSEG03	H41	41110	01/25/2021	9040C
01/25/2021	Water	M1190-08	✓ pH	Cool 4 deg C	PSEG03	H41	41709	01/25/2021	9040C
01/25/2021	Water	M1229-04	✓ pH	Cool 4 deg C	PSEG03	H41	TRE-1136	01/25/2021	9040C
01/20/2021	Water	M1232-01	✓ pH	Cool 4 deg C	STER15	G22	GILROY-OMEGA-CELL	01/20/2021	9040C

Date/Time 01-25-2021 13:40
 Received by: JP (64) (Dem)
 Relinquished by: JP (64) (Dem)

Date/Time 01-25-2021 16:25
 Received by: JP (64)
 Relinquished by: JP (64) (Dem)