



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112911

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/25/2021	15:00
2	CAL2	1	Water	NA	NA	20.2	7.01	01/25/2021	15:02
3	CAL3	1	Water	NA	NA	20.2	10.03	01/25/2021	15:03
4	ICV	1	Water	NA	NA	20.2	6.99	01/25/2021	15:05
5	CCV1	1	Water	NA	NA	20.2	2.01	01/25/2021	15:06
6	M1190-01	1	Solid	20.03	20	20.4	8.15	01/25/2021	15:40
7	M1190-01DUP	1	Solid	20.04	20	20.5	8.16	01/25/2021	15:41
8	M1190-02	1	Solid	20.02	20	20.2	5.78	01/25/2021	15:42
9	M1229-02	1	Solid	20.01	20	20.3	7.20	01/25/2021	15:44
10	M1230-01	1	Solid	20.03	20	20.5	7.40	01/25/2021	15:46
11	CCV2	1	Water	NA	NA	20.2	12.02	01/25/2021	15:48

WORKLIST(Hardcopy Internal Chain)

LB 112911

WorkList Name : ph m1190 s WorkList ID : 145949 Department : Wet-Chemistry Date : 01-25-2021 14:00:47

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/25/2021	Solid	M1190-01	5 pH	Cool 4 deg C	PSEG03	H41	41641	01/25/2021	9045D
01/25/2021	Solid	M1190-02	5 pH	Cool 4 deg C	PSEG03	H41	20796	01/25/2021	9045D
01/25/2021	Solid	M1229-02	10 pH	Cool 4 deg C	PSEG03	H41	TRE-1125	01/25/2021	9045D
01/25/2021	Solid	M1230-01	10 pH	Cool 4 deg C	PSEG03	H41	NWB-1501	01/25/2021	9045D

Date/Time 01-25-2021 14:10
 Received by: JP (wet Chem)
 Relinquished by: JTC (sm)

Date/Time 01-25-2021 16:25
 Received by: JTC (sm)
 Relinquished by: JP (wet Chem)