

LB112913


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

5760E



A= 0.0000e+000

B= 1.7382e-004

C= 1.0135e-002

Rho= 0.9999069

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.022	0.022	68	0.000	68				
0.2	0.200	0.210	0.010	1150	0.0 %	1150				
2.5	2.500	2.430	-0.070	13920	0.0 %	13920				
5.0	5.000	5.075	0.075	29137	0.0 %	29137				
7.5	7.500	7.440	-0.060	42746	0.0 %	42746				
10.0	10.000	10.023	0.023	57606	0.0 %	57606				

0/05
151
-3
2
-1
0

LB112913 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 25 Jan 2021 15:41:07

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	68	-	0.00007471B	PPB		25 Jan 2021 15:42:27	S	Std
0.2 - 1	50.2	1150	-	0.20007471B	PPB		25 Jan 2021 15:44:44	S	Std
2.5 - 1	52.5	13920	-	2.50007471B	PPB		25 Jan 2021 15:47:01	S	Std
5.0 - 1	55	29137	-	5.00007471B	PPB		25 Jan 2021 15:49:18	S	Std
7.5 - 1	57.5	42746	-	7.50007471B	PPB		25 Jan 2021 15:51:35	S	Std
10.0 - 1	510	57606	-	10.00007471B	PPB		25 Jan 2021 15:53:51	S	Std
ICV82 - 1	ICV82	22120	3.8551	-7471B	PPB		25 Jan 2021 15:57:00	U	SMPL
ICB82 - 1	ICB82	-34	0.0042	-7471B	PPB		25 Jan 2021 15:59:15	U	SMPL
CCV75 - 1	CCV75	28882	5.0304	-7471B	PPB		25 Jan 2021 16:01:30	U	SMPL
CCB75 - 1	CCB75	-38	0.0035	-7471B	PPB		25 Jan 2021 16:03:45	U	SMPL
CRA - 1	CRA	1114	0.2038	-7471B	PPB		25 Jan 2021 16:06:00	U	SMPL
HighStd - 1	HighStd	57921	10.0780	-7471B	PPB		25 Jan 2021 16:08:15	U	SMPL
ChkStd - 1	ChkStd	40060	6.9734	-7471B	PPB		25 Jan 2021 16:10:31	U	SMPL
PB134296BL - 1	PBS	-81	-0.0039	-7471B	PPB		25 Jan 2021 16:12:47	U	SMPL
PB134296BS - 1	LCSS	20411	3.5580	-7471B	PPB		25 Jan 2021 16:15:05	U	SMPL
M1189-01 - 1	41108	1065	0.1953	-7471B	PPB		25 Jan 2021 16:17:21	U	SMPL
M1189-02 - 1	41047-41048	5127	0.9013	-7471B	PPB		25 Jan 2021 16:19:38	U	SMPL
M1189-03 - 1	41041	236	0.0512	-7471B	PPB		25 Jan 2021 16:21:55	U	SMPL
M1206-01 - 1	FK-SF-03-009-A	6066	1.0645	-7471B	PPB		25 Jan 2021 16:24:12	U	SMPL
M1206-03 - 1	FK-SF-03-009-B	7447	1.3046	-7471B	PPB		25 Jan 2021 16:26:28	U	SMPL
CCV76 - 1	CCV76	28295	4.9284	-7471B	PPB		25 Jan 2021 16:28:43	U	SMPL
CCB76 - 1	CCB76	-42	0.0028	-7471B	PPB		25 Jan 2021 16:30:58	U	SMPL
M1207-01 - 1	FK-SF-03-010-A	4872	0.8570	-7471B	PPB		25 Jan 2021 16:33:14	U	SMPL
M1207-03 - 1	FK-SF-03-010-B	3960	0.6985	-7471B	PPB		25 Jan 2021 16:35:29	U	SMPL
M1208-01 - 1	FK-SF-03-011-A	5523	0.9701	-7471B	PPB		25 Jan 2021 16:37:45	U	SMPL
M1208-03 - 1	FK-SF-03-011-B	5603	0.9841	-7471B	PPB		25 Jan 2021 16:40:01	U	SMPL
M1209-01 - 1	FK-PS-01-069	3125	0.5533	-7471B	PPB		25 Jan 2021 16:42:17	U	SMPL
M1209-01DUP - 1	FK-PS-01-069DUP	2986	0.5292	-7471B	PPB		25 Jan 2021 16:44:34	U	SMPL
M1209-01MS - 1	FK-PS-01-069MS	23778	4.1432	-7471B	PPB		25 Jan 2021 16:46:50	U	SMPL
M1209-01MSD - 1	FK-PS-01-069MSD	25080	4.3696	-7471B	PPB		25 Jan 2021 16:49:07	U	SMPL
M1215-01 - 1	HR-02-012221	4808	0.8459	-7471B	PPB		25 Jan 2021 16:51:25	U	SMPL
M1220-01 - 1	SP1-A	2974	0.5271	-7471B	PPB		25 Jan 2021 16:53:40	U	SMPL
CCV77 - 1	CCV77	28609	4.9830	-7471B	PPB		25 Jan 2021 16:55:56	U	SMPL
CCB77 - 1	CCB77	-25	0.0058	-7471B	PPB		25 Jan 2021 16:58:12	U	SMPL
M1220-03 - 1	SP1-B	3238	0.5730	-7471B	PPB		25 Jan 2021 17:00:27	U	SMPL
M1220-05 - 1	SP1-C	3216	0.5691	-7471B	PPB		25 Jan 2021 17:02:43	U	SMPL
M1226-06 - 1	0046-051-COMP-H	5831	1.0237	-7471B	PPB		25 Jan 2021 17:04:59	U	SMPL
M1227-01 - 1	20210065-3-2(0.0-0.5)	47549	8.2751	-7471B	PPB		25 Jan 2021 17:07:14	U	SMPL
M1209-01LX5 - 1		368	0.0741	-7471B	PPB		25 Jan 2021 17:09:31	U	SMPL
M1209-01A - 1		28563	4.9750	-7471B	PPB		25 Jan 2021 17:11:49	U	SMPL
CCV78 - 1	CCV78	28911	5.0355	-7471B	PPB		25 Jan 2021 17:14:06	U	SMPL
CCB78 - 1	CCB78	-45	0.0023	-7471B	PPB		25 Jan 2021 17:16:23	U	SMPL