



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB112916

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: FLASHPONT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	28.00	82.4	82.7	01/26/2021	09:40
2	M1190-03		1	9	100.00	>212.0	>212.0	01/26/2021	10:05
3	M1190-03DUP		1	9	100.00	>212.0	>212.0	01/26/2021	10:35
4	M1190-04		1	11	100.00	>212.0	>212.0	01/26/2021	11:10
5	M1190-05		1	10	100.00	>212.0	>212.0	01/26/2021	11:40
6	M1190-06		1	9	100.00	>212.0	>212.0	01/26/2021	12:05
7	M1190-07		1	10	100.00	>212.0	>212.0	01/26/2021	12:32
8	M1190-08		1	11	100.00	>212.0	>212.0	01/26/2021	12:58
9	M1219-01		1	10	100.00	>212.0	>212.0	01/26/2021	13:25
10	M1219-02		1	11	100.00	>212.0	>212.0	01/26/2021	13:50
11	M1229-04		1	10	100.00	>212.0	>212.0	01/26/2021	14:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP12521

WorkList ID : 145952

Department : Wet-Chemistry

Date : 01-25-2021 16:59:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/01/2021	Water	M1190-03	Flash Point	Cool 4 deg C	PSEG03	H41	41109	01/25/2021	1010A
02/01/2021	Water	M1190-04	Flash Point	Cool 4 deg C	PSEG03	H41	41054	01/25/2021	1010A
02/01/2021	Water	M1190-05	Flash Point	Cool 4 deg C	PSEG03	H41	41053	01/25/2021	1010A
02/01/2021	Water	M1190-06	Flash Point	Cool 4 deg C	PSEG03	H41	41111	01/25/2021	1010A
02/01/2021	Water	M1190-07	Flash Point	Cool 4 deg C	PSEG03	H41	41110	01/25/2021	1010A
02/01/2021	Water	M1190-08	Flash Point	Cool 4 deg C	PSEG03	H41	41709	01/25/2021	1010A
01/28/2021	Water	M1219-01	Flash Point	Cool 4 deg C	JACO05	G21	IDW-BR2-AQ-012121-01	01/21/2021	1010A
01/28/2021	Water	M1219-02	Flash Point	Cool 4 deg C	JACO05	G21	IDW-BR2-AQ-012121-02	01/21/2021	1010A
01/28/2021	Water	M1229-04	Flash Point	Cool 4 deg C	PSEG03	H41	TRE-1136	01/25/2021	1010A

Date/Time

Received by:

Relinquished by:

Date/Time

Received by:

Relinquished by: