



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB112917

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: FLASHPONT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	01/25/2021	09:40
2	M1189-06		1	9	100.00	>212.0	>212.0	01/25/2021	10:05
3	M1189-06DUP		1	9	100.00	>212.0	>212.0	01/25/2021	10:35
4	M1189-07		1	10	100.00	>212.0	>212.0	01/25/2021	11:05
5	M1189-08		1	9	100.00	>212.0	>212.0	01/25/2021	11:40
6	M1189-09		1	11	100.00	>212.0	>212.0	01/25/2021	12:15
7	M1189-10		1	10	100.00	>212.0	>212.0	01/25/2021	12:40
8	M1189-11		1	11	100.00	>212.0	>212.0	01/25/2021	13:05
9	M1189-12		1	10	100.00	>212.0	>212.0	01/25/2021	13:35
10	M1192-04		1	9	100.00	>212.0	>212.0	01/25/2021	13:58
11	M1192-05		1	10	100.00	>212.0	>212.0	01/25/2021	14:20
12	M1218-04		1	9	100.00	>212.0	>212.0	01/25/2021	14:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp12121 WorkList ID : 145862 Department : Wet-Chemistry Date : 01-21-2021 09:10:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/28/2021	Water	M1189-06	Flash Point	Cool 4 deg C	PSEG03	H22	41041	01/21/2021	1010A
01/28/2021	Water	M1189-07	Flash Point	Cool 4 deg C	PSEG03	H22	41108	01/21/2021	1010A
01/28/2021	Water	M1189-08	Flash Point	Cool 4 deg C	PSEG03	H22	41051	01/21/2021	1010A
01/28/2021	Water	M1189-09	Flash Point	Cool 4 deg C	PSEG03	H22	41052	01/21/2021	1010A
01/28/2021	Water	M1189-10	Flash Point	Cool 4 deg C	PSEG03	H22	41089-91	01/21/2021	1010A
01/28/2021	Water	M1189-11	Flash Point	Cool 4 deg C	PSEG03	H22	41105-07	01/21/2021	1010A
01/28/2021	Water	M1189-12	Flash Point	Cool 4 deg C	PSEG03	H22	41087-88	01/21/2021	1010A
01/26/2021	Water	M1192-04	Flash Point	Cool 4 deg C	PSEG03	H21	41623	01/21/2021	1010A
01/26/2021	Water	M1192-05	Flash Point	Cool 4 deg C	PSEG03	H21	41046	01/21/2021	1010A
01/29/2021	Water	M1218-04	Flash Point	Cool 4 deg C	PSEG03	H12	238	01/22/2021	1010A

Date/Time 01/25/21 09:15
 Received by: KS (wetchem)
 Relinquished by: JP (wetchem)

Date/Time 01/25/21 14:50
 Received by: JP (wetchem)
 Relinquished by: KS (wetchem)