

LB112919


7470A

INSTRUMENT ID: CV4

Linear

 μ Abs.:

42065



A= 0.0000e+000

B= 2.3605e-004

C= -4.2779e-002

Rho= 0.9997713

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.019	-0.019	102	0.000	102					-
0.2	0.200	0.173	-0.027	916	0.0 %	916					-14
2.5	2.500	2.526	0.026	10884	0.0 %	10884					1
5.0	5.000	4.982	-0.018	21289	0.0 %	21289					0
7.5	7.500	7.650	0.150	32589	0.0 %	32589					2
10.0	10.000	9.887	-0.113	42065	0.0 %	42065					-1

LB112919 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 26 Jan 2021 09:37:09				
Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50	102	-	0.0000	7470A	PPB	26 Jan 2021 09:43:27	S Std
0.2 - 1	50.2	916	-	0.2000	7470A	PPB	26 Jan 2021 09:48:19	S Std
2.5 - 1	50.5	10884	-	2.5000	7470A	PPB	26 Jan 2021 09:50:36	S Std
5.0 - 1	55	21289	-	5.0000	7470A	PPB	26 Jan 2021 09:52:53	S Std
7.5 - 1	57.5	32589	-	7.5000	7470A	PPB	26 Jan 2021 09:55:10	S Std
10.0 - 1	58	42065	-	10.0000	7470A	PPB	26 Jan 2021 09:57:26	S Std
ICV83 - 1	ICV83	15649	3.6511	-	7470A	PPB	26 Jan 2021 10:00:15	U SMPL
ICB83 - 1	ICB83	37	-0.0340	-	7470A	PPB	26 Jan 2021 10:02:30	U SMPL
CCV79 - 1	CCV79	21337	4.9938	-	7470A	PPB	26 Jan 2021 10:04:45	U SMPL
CCB79 - 1	CCB79	122	-0.0140	-	7470A	PPB	26 Jan 2021 10:06:59	U SMPL
CRA - 1	CRA	910	0.1720	-	7470A	PPB	26 Jan 2021 10:09:14	U SMPL
HighStd - 1	HighStd	41664	9.7920	-	7470A	PPB	26 Jan 2021 10:11:30	U SMPL
ChkStd - 1	ChkStd	28149	6.6018	-	7470A	PPB	26 Jan 2021 10:13:45	U SMPL
PB134304BL - 1	PBW	173	-0.0019	-	7470A	PPB	26 Jan 2021 10:16:01	U SMPL
PB134304BS - 1	LCSW	15572	3.6330	-	7470A	PPB	26 Jan 2021 10:20:54	U SMPL
M1119-02 - 1	IDW-BR1-SO-011221	38	-0.0338	-	7470A	PPB	26 Jan 2021 10:23:10	U SMPL
M1188-03 - 1	TP-230	553	0.0878	-	7470A	PPB	26 Jan 2021 10:25:26	U SMPL
M1188-06 - 1	TP-SWITCH-GEAR	596	0.0979	-	7470A	PPB	26 Jan 2021 10:27:43	U SMPL
M1192-01 - 1	41015	580	0.0941	-	7470A	PPB	26 Jan 2021 10:30:00	U SMPL
M1192-02 - 1	41016	626	0.1050	-	7470A	PPB	26 Jan 2021 10:32:16	U SMPL
CCV80 - 1	CCV80	21422	5.0139	-	7470A	PPB	26 Jan 2021 10:34:31	U SMPL
CCB80 - 1	CCB80	120	-0.0145	-	7470A	PPB	26 Jan 2021 10:36:47	U SMPL
M1192-03 - 1	41013	549	0.0868	-	7470A	PPB	26 Jan 2021 10:39:02	U SMPL
M1213-02 - 1	IDW-BR2-SO-012121	14647	3.4146	-	7470A	PPB	26 Jan 2021 10:41:18	U SMPL
M1213-02DUP - 1	IDW-BR2-SO-012121DUP	13158	3.0631	-	7470A	PPB	26 Jan 2021 10:43:33	U SMPL
M1213-02MS - 1	IDW-BR2-SO-012121MS	28936	6.7875	-	7470A	PPB	26 Jan 2021 10:51:00	U SMPL
M1213-02MSD - 1	IDW-BR2-SO-012121MSD	29390	6.8947	-	7470A	PPB	26 Jan 2021 10:53:16	U SMPL
M1216-01 - 1	PS-105	47	-0.0317	-	7470A	PPB	26 Jan 2021 10:56:11	U SMPL
M1216-02 - 1	PS-104	567	0.0911	-	7470A	PPB	26 Jan 2021 10:58:27	U SMPL
M1216-03 - 1	PS-102	575	0.0929	-	7470A	PPB	26 Jan 2021 11:00:44	U SMPL
M1217-01 - 1	1617	659	0.1128	-	7470A	PPB	26 Jan 2021 11:03:01	U SMPL
M1218-01 - 1	240	672	0.1158	-	7470A	PPB	26 Jan 2021 11:05:16	U SMPL
CCV81 - 1	CCV81	22375	5.2388	-	7470A	PPB	26 Jan 2021 11:07:32	U SMPL
CCB81 - 1	CCB81	112	-0.0163	-	7470A	PPB	26 Jan 2021 11:09:48	U SMPL
M1218-03 - 1	241	643	0.1090	-	7470A	PPB	26 Jan 2021 11:12:03	U SMPL
PB134305BL - 1	PBW	45	-0.0322	-	7470A	PPB	26 Jan 2021 11:16:53	U SMPL
PB134305BS - 1	LCSW	14345	3.3433	-	7470A	PPB	26 Jan 2021 11:19:08	U SMPL
M1216-04 - 1	PS-103	600	0.0988	-	7470A	PPB	26 Jan 2021 11:21:23	U SMPL
M1217-15 - 1	1525	569	0.0915	-	7470A	PPB	26 Jan 2021 11:23:39	U SMPL
M1217-16 - 1	1710	514	0.0785	-	7470A	PPB	26 Jan 2021 11:25:55	U SMPL
M1217-17 - 1	1723-1724	530	0.0823	-	7470A	PPB	26 Jan 2021 11:28:12	U SMPL
M1217-18 - 1	1726-1727	476	0.0696	-	7470A	PPB	26 Jan 2021 11:30:29	U SMPL
M1217-19 - 1	1674-1685	537	0.0840	-	7470A	PPB	26 Jan 2021 11:32:46	U SMPL
M1217-19DUP - 1	1674-1685DUP	573	0.0925	-	7470A	PPB	26 Jan 2021 11:35:02	U SMPL
CCV82 - 1	CCV82	21108	4.9397	-	7470A	PPB	26 Jan 2021 11:37:18	U SMPL
CCB82 - 1	CCB82	113	-0.0161	-	7470A	PPB	26 Jan 2021 11:39:34	U SMPL
M1217-19MS - 1	1674-1685MS	14723	3.4326	-	7470A	PPB	26 Jan 2021 11:41:49	U SMPL
M1217-19MSD - 1	1674-1685MSD	15483	3.6120	-	7470A	PPB	26 Jan 2021 11:44:05	U SMPL
PB134306BL - 1	PBW	42	-0.0329	-	7470A	PPB	26 Jan 2021 11:49:00	U SMPL
PB134306BS - 1	LCSW	15790	3.6844	-	7470A	PPB	26 Jan 2021 11:56:14	U SMPL
M1218-04 - 1	238	414	0.0549	-	7470A	PPB	26 Jan 2021 11:58:30	U SMPL
M1218-06 - 1	237	383	0.0476	-	7470A	PPB	26 Jan 2021 12:00:46	U SMPL
M1226-01 - 1	0049-043-COMP-C-ELUTRIATE	740	0.1319	-	7470A	PPB	26 Jan 2021 12:03:03	U SMPL
M1226-01DUP - 1	0049-043-COMP-C-ELUTRIATEDUP	666	0.1144	-	7470A	PPB	26 Jan 2021 12:05:20	U SMPL
M1226-01MS - 1	0049-043-COMP-C-ELUTRIATEMS	14571	3.3967	-	7470A	PPB	26 Jan 2021 12:07:38	U SMPL
M1226-01MSD - 1	0049-043-COMP-C-ELUTRIATEMSD	14903	3.4751	-	7470A	PPB	26 Jan 2021 12:09:54	U SMPL
CCV83 - 1	CCV83	21053	4.9268	-	7470A	PPB	26 Jan 2021 12:12:10	U SMPL
CCB83 - 1	CCB83	193	0.0028	-	7470A	PPB	26 Jan 2021 12:14:27	U SMPL
M1226-02 - 1	0050-044-COMP-D-ELUTRIATE	635	0.1071	-	7470A	PPB	26 Jan 2021 12:16:42	U SMPL
M1226-03 - 1	0051-045-COMP-E-ELUTRIATE	522	0.0804	-	7470A	PPB	26 Jan 2021 12:18:58	U SMPL
M1226-04 - 1	0055-052-COMP-I-ELUTRIATE	520	0.0800	-	7470A	PPB	26 Jan 2021 12:21:14	U SMPL
M1226-05 - 1	0049-043-COMP-C-ELUTRIATE	207	0.0061	-	7470A	PPB	26 Jan 2021 12:23:29	U SMPL
M1226-07 - 1	0050-044-COMP-D-ELUTRIATE	208	0.0063	-	7470A	PPB	26 Jan 2021 12:25:45	U SMPL
M1226-08 - 1	0051-045-COMP-E-ELUTRIATE	135	-0.0109	-	7470A	PPB	26 Jan 2021 12:28:02	U SMPL
M1226-09 - 1	0055-052-COMP-I-ELUTRIATE	143	-0.0090	-	7470A	PPB	26 Jan 2021 12:30:18	U SMPL
PB134304TB - 1	PB134269TB	437	0.0604	-	7470A	PPB	26 Jan 2021 12:32:36	U SMPL
PB134305TB - 1	PB134295TB	391	0.0495	-	7470A	PPB	26 Jan 2021 12:34:53	U SMPL
CCV84 - 1	CCV84	21478	5.0271	-	7470A	PPB	26 Jan 2021 12:37:08	U SMPL
CCB84 - 1	CCB84	-24	-0.0484	-	7470A	PPB	26 Jan 2021 12:39:24	U SMPL
M1213-02LX5 - 1		3021	0.6703	-	7470A	PPB	26 Jan 2021 12:41:40	U SMPL
M1213-02A - 1		29290	6.8711	-	7470A	PPB	26 Jan 2021 12:43:57	U SMPL
M1217-19LX5 - 1		30	-0.0357	-	7470A	PPB	26 Jan 2021 12:46:13	U SMPL
M1217-19A - 1		15769	3.6795	-	7470A	PPB	26 Jan 2021 12:48:29	U SMPL
M1226-01LX5 - 1		97	-0.0199	-	7470A	PPB	26 Jan 2021 12:50:46	U SMPL
M1226-01A - 1		17445	4.0751	-	7470A	PPB	26 Jan 2021 12:53:03	U SMPL
CCV85 - 1	CCV85	21794	5.1017	-	7470A	PPB	26 Jan 2021 12:55:19	U SMPL
CCB85 - 1	CCB85	149	-0.0076	-	7470A	PPB	26 Jan 2021 12:57:37	U SMPL