



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB112925

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	01/26/2021	13:25
2	CAL2	1	Water	NA	NA	21.2	7.01	01/26/2021	13:26
3	CAL3	1	Water	NA	NA	21.2	10.04	01/26/2021	13:27
4	ICV	1	Water	NA	NA	21.2	6.98	01/26/2021	13:30
5	CCV1	1	Water	NA	NA	21.3	2.01	01/26/2021	13:32
6	M1234-02	1	Solid	20.03	20	20.2	9.24	01/26/2021	13:34
7	M1239-01	1	Solid	20.03	20	20.6	5.33	01/26/2021	13:35
8	M1240-01	1	Solid	20.01	20	20.2	5.35	01/26/2021	13:37
9	M1240-03	1	Solid	20.04	20	20.3	6.66	01/26/2021	13:40
10	M1240-03DUP	1	Solid	20.04	20	20.4	6.67	01/26/2021	13:41
11	CCV2	1	Water	NA	NA	20.3	12.02	01/26/2021	13:42

WORKLIST(Hardcopy Internal Chain)

13112925

WorkList Name : CORROSIONITY M1239 WorkList ID : 145989 Department : Wet-Chemistry Date : 01-26-2021 13:26:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/25/2021	Solid	M1234-02	Corrosivity	Cool 4 deg C	TULL02	G22	HOLLAND TUNNEL-TEST-PI	01/25/2021	9045D
01/26/2021	Solid	M1239-01	Corrosivity	Cool 4 deg C	PSEG03	H51	N36712	01/26/2021	9045D
01/26/2021	Solid	M1240-01	Corrosivity	Cool 4 deg C	PSEG03	H51	72-11935	01/26/2021	9045D
01/26/2021	Solid	M1240-03	Corrosivity	Cool 4 deg C	PSEG03	H51	72-11995	01/26/2021	9045D

Date/Time 01-26-2021 13:20
 Received by: JO (Ext. Lab)
 Relinquished by: AP (Ext. Lab)

Date/Time 01-26-2021 15:00
 Received by: AP (Ext. Lab)
 Relinquished by: JO (Ext. Lab)