

LB112926


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

45140



A= 0.0000e+000

B= 2.2193e-004 *slope*C= -1.1408e-002 *Intercept*

Rho= 0.9999914

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>0/0 B</i>
0.00	0.000	-0.002	-0.002	43	0.000	43					-
0.05	0.050	0.063	0.013	334	0.0 %	334					26
0.20	0.200	0.208	0.008	989	0.0 %	989					4
2.50	2.500	2.492	-0.008	11280	0.0 %	11280					0
5.00	5.000	4.967	-0.033	22434	0.0 %	22434					-1
7.50	7.500	7.516	0.016	33917	0.0 %	33917					0
10.0	10.000	10.006	0.006	45140	0.0 %	45140					0

LB112926
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0.05		43	-		0 ISM02.4	PPB	1/26/2021 14:03	Std
0.05 S0.05		334	-	0.05	ISM02.4	PPB	1/26/2021 14:05	Std
0.2 S0.2		989	-	0.2	ISM02.4	PPB	1/26/2021 14:07	Std
2.5 S2.5		11280	-	2.5	ISM02.4	PPB	1/26/2021 14:09	Std
5 S5.0		22434	-	5	ISM02.4	PPB	1/26/2021 14:12	Std
7.5 S7.5		33917	-	7.5	ISM02.4	PPB	1/26/2021 14:14	Std
10 S10.0		45140	-	10	ISM02.4	PPB	1/26/2021 14:27	Std
ICV	ICV	17440	3.859 -		ISM02.4	PPB	1/26/2021 14:30	SMPL
ICB	ICB	0	-0.0114 -		ISM02.4	PPB	1/26/2021 14:32	SMPL
CCV086	CCV086	22696	5.0254 -		ISM02.4	PPB	1/26/2021 14:34	SMPL
CCB086	CCB086	-6	-0.0127 -		ISM02.4	PPB	1/26/2021 14:37	SMPL
PB134341BL	PBW341	35	-0.0036 -		ISM02.4	PPB	1/26/2021 14:39	SMPL
M1183-01	MEW0Z7	41	-0.0023 -		ISM02.4	PPB	1/26/2021 14:41	SMPL
M1183-02	MEW0Z8	11	-0.009 -		ISM02.4	PPB	1/26/2021 14:43	SMPL
M1183-03	MEW100	30	-0.0048 -		ISM02.4	PPB	1/26/2021 14:46	SMPL
M1183-04	MEW100D	26	-0.0056 -		ISM02.4	PPB	1/26/2021 14:48	SMPL
M1183-05	MEW100S	4303	0.9435 -		ISM02.4	PPB	1/26/2021 14:50	SMPL
CCV087	CCV087	22610	5.0063 -		ISM02.4	PPB	1/26/2021 14:53	SMPL
CCB087	CCB087	4	-0.0105 -		ISM02.4	PPB	1/26/2021 14:55	SMPL