

LB 112928


SFAM01.1

INSTRUMENT ID: CX1

Linear

 μ Abs.:

45056



A= 0.0000e+000

B= 2.2211e-004 slope

C= -1.9353e-002 y-intercept

Rho= 0.9999929

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.004	-0.004	67	0.000	67					
0.05	0.050					334					
0.20	0.200	0.201	0.001	994	0.0 %	994					
2.50	2.500	2.508	0.008	11378	0.0 %	11378					
5.00	5.000	4.982	-0.018	22518	0.0 %	22518					
7.50	7.500	7.525	0.025	33966	0.0 %	33966					
10.0	10.000	9.988	-0.012	45056	0.0 %	45056					

10000

LB112928
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	67	-		0 SFAM01.1	PPB	1/26/2021 15:39	Std
	0.2 S01	994	-		0.2 SFAM01.1	PPB	1/26/2021 15:41	Std
	2.5 S02	11378	-		2.5 SFAM01.1	PPB	1/26/2021 15:44	Std
	5 S03	22518	-		5 SFAM01.1	PPB	1/26/2021 15:46	Std
	7.5 S04	33966	-		7.5 SFAM01.1	PPB	1/26/2021 15:48	Std
	10 S05	45056	-		10 SFAM01.1	PPB	1/26/2021 15:50	Std
ICV010	ICV010	17516	3.8712	-	SFAM01.1	PPB	1/26/2021 15:53	SMPL
ICB010	ICB010	-15	-0.0227	-	SFAM01.1	PPB	1/26/2021 15:55	SMPL
CCV088	CCV088	22725	5.0281	-	SFAM01.1	PPB	1/26/2021 15:58	SMPL
CCB088	CCB088	8	-0.0176	-	SFAM01.1	PPB	1/26/2021 16:00	SMPL
PB134342BL	PBW342	38	-0.0109	-	SFAM01.1	PPB	1/26/2021 16:02	SMPL
PB134342TB	LEB342	119	0.0071	-	SFAM01.1	PPB	1/26/2021 16:04	SMPL
M1205-01	X2X98	732786	HIGH	-	SFAM01.1	PPB	1/26/2021 16:07	
M1205-02	X2X98D	722840	HIGH	-	SFAM01.1	PPB	1/26/2021 16:08	
M1205-03	X2X98S	761296	HIGH	-	SFAM01.1	PPB	1/26/2021 16:10	
CCV089	CCV089	23091	5.1094	-	SFAM01.1	PPB	1/26/2021 16:12	SMPL
CCB089	CCB089	-218	-0.0678	-	SFAM01.1	PPB	1/26/2021 16:15	SMPL
M1205-01DLX10	X2X98	150414	33.3894	-	SFAM01.1	PPB	1/26/2021 16:17	SMPL
M1205-02DLX10	X2X98D	147826	32.8146	-	SFAM01.1	PPB	1/26/2021 16:19	SMPL
M1205-03DLX10	X2X98S	161167	35.7778	-	SFAM01.1	PPB	1/26/2021 16:22	SMPL
M1205-01DL2X50	X2X98	29788	6.5969	-	SFAM01.1	PPB	1/26/2021 16:24	SMPL
M1205-02DL2X50	X2X98D	30735	6.8073	-	SFAM01.1	PPB	1/26/2021 16:26	SMPL
M1205-03DL2X50	X2X98S	32861	7.2795	-	SFAM01.1	PPB	1/26/2021 16:29	SMPL
CCV090	CCV090	22459	4.9691	-	SFAM01.1	PPB	1/26/2021 16:31	SMPL
CCB090	CCB090	19	-0.0151	-	SFAM01.1	PPB	1/26/2021 16:33	SMPL