


LB112932

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.4497e-004

C= 2.3602e-002

Rho= 0.9994531

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%b
0.0	0.000	0.021	0.021	-9	0.000	-9					-
0.2	0.200	0.202	0.002	728	0.0 %	728					1
2.5	2.500	2.513	0.013	10162	0.0 %	10162					1
5.0	5.000	5.078	0.078	20632	0.0 %	20632					2
7.5	7.500	7.248	-0.252	29490	0.0 %	29490					-3
10.0	10.000	10.138	0.138	41288	0.0 %	41288					1

LB112932 IN STRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Jan 2021 10:02:56

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-9	-	0.00007470A	PPB	27 Jan 2021 10:05:34	S	Std
0.2 - 1	50.2	728	-	0.20007470A	PPB	27 Jan 2021 10:07:51	S	Std
2.5 - 1	52.5	10162	-	2.50007470A	PPB	27 Jan 2021 10:10:08	S	Std
5.0 - 1	55	20632	-	5.00007470A	PPB	27 Jan 2021 10:12:25	S	Std
7.5 - 1	57.5	29490	-	7.50007470A	PPB	27 Jan 2021 10:14:42	S	Std
10.0 - 1	59.5	41288	-	10.00007470A	PPB	27 Jan 2021 10:19:32	S	Std
ICV84 - 1	ICV84	15913	3.9218	-7470A	PPB	27 Jan 2021 10:22:10	U	SMPL
ICB84 - 1	ICB84	-67	0.0072	-7470A	PPB	27 Jan 2021 10:24:25	U	SMPL
CCV90 - 1	CCV90	19823	4.8797	-7470A	PPB	27 Jan 2021 10:26:40	U	SMPL
CCB90 - 1	CCB90	-65	0.0077	-7470A	PPB	27 Jan 2021 10:28:55	U	SMPL
CRA - 1	CRA	748	0.2068	-7470A	PPB	27 Jan 2021 10:31:11	U	SMPL
HighStd - 1	HighStd	41035	10.0760	-7470A	PPB	27 Jan 2021 10:33:26	U	SMPL
ChkStd - 1	ChkStd	28536	7.0141	-7470A	PPB	27 Jan 2021 10:38:14	U	SMPL
PB134347BL - 1	PBW	-101	-0.0011	-7470A	PPB	27 Jan 2021 10:40:29	U	SMPL
PB134347BS - 1	LCSW	14094	3.4762	-7470A	PPB	27 Jan 2021 10:42:46	U	SMPL
M1229-01 - 1	TRE-1125	-53	0.0106	-7470A	PPB	27 Jan 2021 10:45:02	U	SMPL
M1234-02 - 1	HOLLAND TUNNEL-TEST-PIT-5	5	0.0248	-7470A	PPB	27 Jan 2021 10:47:18	U	SMPL
M1234-02DUP - 1	HOLLAND TUNNEL-TEST-PIT-5DUP	22	0.0290	-7470A	PPB	27 Jan 2021 10:49:36	U	SMPL
M1234-02MS - 1	HOLLAND TUNNEL-TEST-PIT-5MS	15223	3.7528	-7470A	PPB	27 Jan 2021 10:51:52	U	SMPL
M1234-02MSD - 1	HOLLAND TUNNEL-TEST-PIT-5MSD	16116	3.9716	-7470A	PPB	27 Jan 2021 10:54:08	U	SMPL
CCV91 - 1	CCV91	20109	4.9497	-7470A	PPB	27 Jan 2021 10:56:23	U	SMPL
CCB91 - 1	CCB91	-78	0.0045	-7470A	PPB	27 Jan 2021 10:58:38	U	SMPL
PB134347TB - 1	PB134335TB	-10	0.0212	-7470A	PPB	27 Jan 2021 11:00:54	U	SMPL
M1234-02LX5 - 1		3	0.0243	-7470A	PPB	27 Jan 2021 11:03:09	U	SMPL
M1234-02A - 1		16053	3.9561	-7470A	PPB	27 Jan 2021 11:05:25	U	SMPL
CCV92 - 1	CCV92	20217	4.9762	-7470A	PPB	27 Jan 2021 11:07:41	U	SMPL
CCB92 - 1	CCB92	-40	0.0138	-7470A	PPB	27 Jan 2021 11:09:57	U	SMPL