

LB112940


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

56264



A= 0.0000e+000

B= 1.7656e-004

C= 2.6404e-002

Rho= 0.9998647

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/b
0.0	0.000	0.035	0.035	51	0.000	51					1
0.2	0.200	0.222	0.022	1109	0.0 %	1109					1
2.5	2.500	2.420	-0.080	13554	0.0 %	13554					3
5.0	5.000	4.959	-0.041	27938	0.0 %	27938					1
7.5	7.500	7.603	0.103	42913	0.0 %	42913					1
10.0	10.000	9.960	-0.040	56264	0.0 %	56264					0

LB112940 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 27 Jan 2021 14:15:00				
Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	SD	51	-	0.00007471B	PPB	27 Jan 2021 14:16:01	S	Std
0.2 - 1	SD2	1109	-	0.20007471B	PPB	27 Jan 2021 14:18:18	S	Std
2.5 - 1	SD5	13554	-	2.50007471B	PPB	27 Jan 2021 14:20:35	S	Std
5.0 - 1	SD5	27938	-	5.00007471B	PPB	27 Jan 2021 14:22:52	S	Std
7.5 - 1	SD5	42913	-	7.50007471B	PPB	27 Jan 2021 14:25:09	S	Std
10.0 - 1	SD10	56264	-	10.00007471B	PPB	27 Jan 2021 14:27:25	S	Std
ICV85 - 1	ICV85	22027	3.9155	-7471B	PPB	27 Jan 2021 14:30:08	U	SMPL
ICB85 - 1	ICB85	-55	0.0167	-7471B	PPB	27 Jan 2021 14:32:23	U	SMPL
CCV95 - 1	CCV95	27821	4.9385	-7471B	PPB	27 Jan 2021 14:34:38	U	SMPL
CCB95 - 1	CCB95	-82	0.0119	-7471B	PPB	27 Jan 2021 14:36:54	U	SMPL
CRA - 1	CRA	1102	0.2210	-7471B	PPB	27 Jan 2021 14:39:09	U	SMPL
HighStd - 1	HighStd	55762	9.8718	-7471B	PPB	27 Jan 2021 14:41:25	U	SMPL
ChkStd - 1	ChkStd	40339	7.1487	-7471B	PPB	27 Jan 2021 14:43:40	U	SMPL
PB134346BL - 1	PBS	-101	0.0086	-7471B	PPB	27 Jan 2021 14:45:57	U	SMPL
PB134346BS - 1	LCSS	21773	3.8707	-7471B	PPB	27 Jan 2021 14:48:15	U	SMPL
M1190-01 - 1	41641	706	0.1511	-7471B	PPB	27 Jan 2021 14:50:31	U	SMPL
M1190-01DUP - 1	41641DUP	640	0.1394	-7471B	PPB	27 Jan 2021 14:52:48	U	SMPL
M1190-01MS - 1	41641MS	22187	3.9438	-7471B	PPB	27 Jan 2021 14:55:06	U	SMPL
M1190-01MSD - 1	41641MSD	22599	4.0165	-7471B	PPB	27 Jan 2021 14:57:23	U	SMPL
M1190-02 - 1	20796	536	0.1210	-7471B	PPB	27 Jan 2021 14:59:38	U	SMPL
CCV96 - 1	CCV96	28318	5.0263	-7471B	PPB	27 Jan 2021 15:01:54	U	SMPL
CCB96 - 1	CCB96	-40	0.0193	-7471B	PPB	27 Jan 2021 15:04:09	U	SMPL
M1228-01 - 1	OK-01-012521	1224	0.2425	-7471B	PPB	27 Jan 2021 15:06:25	U	SMPL
M1230-01 - 1	NWB-1501	753	0.1594	-7471B	PPB	27 Jan 2021 15:08:40	U	SMPL
M1231-01 - 1	EO-02-012521	3016	0.5589	-7471B	PPB	27 Jan 2021 15:10:56	U	SMPL
M1234-01 - 1	HOLLAND TUNNEL-TEST-PIT-5	64510	11.4164	-7471B	PPB	27 Jan 2021 15:13:12	U	SMPL
M1235-01 - 1	FK-SF-SAMPLE-1	5125	0.9313	-7471B	PPB	27 Jan 2021 15:15:27	U	SMPL
M1235-03 - 1	FK-SF-SAMPLE-2	6018	1.0890	-7471B	PPB	27 Jan 2021 15:17:45	U	SMPL
M1236-02 - 1	L229R-WO(S)-1-14-1	1182	0.2351	-7471B	PPB	27 Jan 2021 15:20:02	U	SMPL
M1236-04 - 1	J2288R-WO(N)-1-13-1	575	0.1279	-7471B	PPB	27 Jan 2021 15:22:19	U	SMPL
M1236-06 - 1	2922	2450	0.4590	-7471B	PPB	27 Jan 2021 15:24:36	U	SMPL
M1236-08 - 1	PS1419LN	558	0.1249	-7471B	PPB	27 Jan 2021 15:26:52	U	SMPL
CCV97 - 1	CCV97	28140	4.9949	-7471B	PPB	27 Jan 2021 15:29:08	U	SMPL
CCB97 - 1	CCB97	-24	0.0222	-7471B	PPB	27 Jan 2021 15:31:24	U	SMPL
M1237-01 - 1	SU-01-012621	2673	0.4984	-7471B	PPB	27 Jan 2021 15:33:40	U	SMPL
M1238-01 - 1	HD-02-012621	2621	0.4892	-7471B	PPB	27 Jan 2021 15:35:55	U	SMPL
M1241-01 - 1	FACILITY-1	29054	5.1562	-7471B	PPB	27 Jan 2021 15:38:10	U	SMPL
M1242-01 - 1	0014-048-COMP-G	3370	0.6214	-7471B	PPB	27 Jan 2021 15:40:26	U	SMPL
M1190-01LX5 - 1		87	0.0418	-7471B	PPB	27 Jan 2021 15:42:42	U	SMPL
M1190-01A - 1		24655	4.3795	-7471B	PPB	27 Jan 2021 15:44:58	U	SMPL
M1234-01DLX2 - 1		32560	5.7753	-7471B	PPB	27 Jan 2021 15:47:15	U	SMPL
CCV98 - 1	CCV98	28176	5.0012	-7471B	PPB	27 Jan 2021 15:49:32	U	SMPL
CCB98 - 1	CCB98	-14	0.0239	-7471B	PPB	27 Jan 2021 15:51:50	U	SMPL