



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 1/29/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 17:20
In Date: 01/27/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 01/28/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB112941

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
M1089-11	DUP-15	1	1.08	9.58	7.71	78.0	
M1089-11D	DUP-15DUP	2	1.06	9.51	7.73	78.9	
M1250-01	T1-CONTAINMENT	11	1.00	2.00	2.00	100.0	stone sample
M1250-02	T1-CONTAINMENT	12	1.00	2.00	2.00	100.0	stone sample
M1251-01	400026	13	1.11	9.56	7.12	71.1	
M1251-02	400026	14	1.11	9.56	7.03	70.1	
M1251-03	200010	15	1.18	9.64	7.78	78.0	
M1251-04	200010	16	1.18	9.64	7.8	78.3	
M1251-05	200059	17	1.17	9.64	7.2	71.2	
M1251-06	200059	18	1.17	9.64	7.13	70.4	
M1251-07	400024	19	1.18	9.85	7.14	68.7	
M1251-08	400024	20	1.18	9.85	7.03	67.5	
M1252-01	BS-SP	21	1.17	9.83	8.64	86.3	
M1252-02	BS-SP-EPH-2	22	1.17	9.55	8.31	85.2	
M1252-03	BS-CON	23	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M1253-01	124055	24	1.16	9.54	8.95	93.0	
M1253-02	124055-EPH-2	25	1.17	9.74	9.18	93.5	
M1253-03	124035	26	1.15	9.6	8.39	85.7	
M1253-04	124035-EPH-2	27	1.17	9.66	8.53	86.7	
M1253-05	VNJ-201	28	1.16	9.67	8.23	83.1	
M1253-06	VNJ-201-EPH-2	29	1.16	9.83	8.28	82.1	
M1253-07	VNJ-229	30	1.15	9.83	8.36	83.1	
M1253-08	VNJ-229-EPH-2	31	1.13	9.62	8.19	83.2	
M1254-01	CRT-15	32	1.16	9.61	7.61	76.3	
M1254-05	CRT-15-EPH-2	33	1.16	9.51	6.55	64.6	
M1254-06	CRT-5	34	1.14	9.94	9.33	93.1	
M1254-10	CRT-5-EPH-2	35	1.15	9.71	9.13	93.2	
M1255-01	CRT-3	36	1.17	9.74	8.32	83.4	



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 1/29/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 17:20
In Date: 01/27/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 01/28/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB112941

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M1255-05	CRT-3-EPH-2	37	1.18	9.89	8.52	84.3	
M1255-06	CRT-13	38	1.17	9.84	8.52	84.8	
M1255-10	CRT-13-EPH-2	39	1.18	9.58	8.15	83.0	
M1255-11	CRT-17	40	1.16	9.75	8.96	90.8	
M1255-15	CRT-17-EPH-2	41	1.18	9.92	9.11	90.7	
M1256-01	SIDEWALK	55	1.18	9.71	8.37	84.3	
M1256-02	SIDEWALK	56	1.18	9.8	8.42	84.0	
M1256-03	SIDEWALK-EPH-2	57	1.17	9.64	8.44	85.8	
M1257-01	72-11938-1	42	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M1257-03	72-11938-2-POLES	43	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M1260-03	WOOSPK-069-0002	3	1.08	9.75	8.99	91.2	
M1260-04	WOOSPK-069-0204	4	1.08	9.67	7.56	75.4	
M1260-13	WOOSPK-054-0002	5	1.08	9.98	9.08	89.9	
M1260-14	WOOSPK-054-0204	6	1.11	9.86	7.39	71.8	
M1260-20	WOOSPK-057-0204	7	1.08	9.84	6.96	67.1	
M1260-29	2021-01-26-FD02	8	1.09	9.74	7.64	75.7	
M1260-31	M1260-20MS	9	1.08	9.84	6.96	67.1	
M1260-32	M1260-20MSD	10	1.08	9.84	6.96	67.1	
M1262-01	JTY2-WC-S-1-1.0-6.0	44	1.17	9.63	7.35	73.0	
M1264-01	S1-01-E	45	1.16	9.71	7.58	75.1	
M1264-02	S1-02-E	46	1.17	9.58	7.65	77.1	
M1264-03	S1-03-E	47	1.18	9.8	7.57	74.1	
M1264-04	S1-04-E	48	1.16	9.8	7.7	75.7	
M1264-05	S1-05-E	49	1.17	9.78	7.98	79.1	
M1264-06	S1-06-E	50	1.17	9.6	7.34	73.2	
M1264-13	DUP-20	51	1.18	9.51	7.94	81.2	
M1265-01	S3-01-E	52	1.18	9.66	7.89	79.1	
M1265-02	S3-02-E	53	1.18	9.86	7.7	75.1	



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 1/29/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 17:20
In Date: 01/27/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 01/28/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB112941

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M1265-03	S3-03-E	54	1.18	9.56	7.76	78.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$