



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB112950

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	01/28/2021	10:48
2	CAL2	1	Water	NA	NA	20.2	7.00	01/28/2021	10:49
3	CAL3	1	Water	NA	NA	20.2	10.03	01/28/2021	10:50
4	ICV	1	Water	NA	NA	20.2	7.00	01/28/2021	10:51
5	CCV1	1	Water	NA	NA	20.2	2.01	01/28/2021	10:52
6	M1174-02	1	Solid	20.03	20	20.3	6.42	01/28/2021	10:55
7	M1236-02	1	Solid	20.04	20	20.3	8.72	01/28/2021	10:57
8	M1236-04	1	Solid	20.01	20	20.4	8.03	01/28/2021	10:59
9	M1236-06	1	Solid	20.03	20	20.4	8.15	01/28/2021	11:00
10	M1236-08	1	Solid	20.04	20	20.6	8.64	01/28/2021	11:02
11	M1254-02	1	Solid	20.02	20	20.1	7.77	01/28/2021	11:05
12	M1254-07	1	Solid	20.04	20	20.4	8.07	01/28/2021	11:06
13	M1255-02	1	Solid	20.03	20	20.6	7.03	01/28/2021	11:08
14	M1255-07	1	Solid	20.04	20	20.4	8.07	01/28/2021	11:10
15	M1255-12	1	Solid	20.03	20	20.5	7.23	01/28/2021	11:13
16	CCV2	1	Water	NA	NA	20.2	12.02	01/28/2021	11:15
17	M1262-01	1	Solid	20.04	20	20.3	8.77	01/28/2021	11:18
18	M1262-01DUP	1	Solid	20.03	20	20.3	8.78	01/28/2021	11:19
19	CCV3	1	Water	NA	NA	20.3	2.01	01/28/2021	11:20

WORKLIST(Hardcopy Internal Chain)

112950

WorkList Name : CORROSIONITY M1262 WorkList ID : 146034 Department : Wet-Chemistry Date : 01-28-2021 10:40:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/27/2021	Solid	M1174-02	Corrosivity	Cool 4 deg C	CLEA11	G22	OWS-1	01/27/2021	9045D
01/26/2021	Solid	M1236-02	Corrosivity	Cool 4 deg C	PSEG03	H51	L229R-WO(S)-1-14-1	01/26/2021	9045D
01/26/2021	Solid	M1236-04	Corrosivity	Cool 4 deg C	PSEG03	H51	J2288R-WO(N)-1-13-1	01/26/2021	9045D
01/26/2021	Solid	M1236-06	Corrosivity	Cool 4 deg C	PSEG03	H51	2922	01/26/2021	9045D
01/26/2021	Solid	M1236-08	Corrosivity	Cool 4 deg C	PSEG03	H51	PS1419LN	01/26/2021	9045D
01/27/2021	Solid	M1254-02	Corrosivity	Cool 4 deg C	PSEG03	H61	CRT-15	01/27/2021	9045D
01/27/2021	Solid	M1254-07	Corrosivity	Cool 4 deg C	PSEG03	H61	CRT-5	01/27/2021	9045D
01/27/2021	Solid	M1255-02	Corrosivity	Cool 4 deg C	PSEG03	H63	CRT-3	01/27/2021	9045D
01/27/2021	Solid	M1255-07	Corrosivity	Cool 4 deg C	PSEG03	H63	CRT-13	01/27/2021	9045D
01/27/2021	Solid	M1255-12	Corrosivity	Cool 4 deg C	PSEG03	H63	CRT-17	01/27/2021	9045D
01/26/2021	Solid	M1262-01	Corrosivity	Cool 4 deg C	PARS15	G32	JTY2-WC-S-1-1.0-6.0	01/26/2021	9045D

Date/Time 01/28/2021 10:45
 Received by: JO wet chem
 Relinquished by: JO wet chem

Date/Time 01/28/2021 12:10
 Received by: JO wet chem
 Relinquished by: JO wet chem