



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112951

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/28/2021	11:00
2	CAL2	1	Water	NA	NA	20.1	7.00	01/28/2021	11:01
3	CAL3	1	Water	NA	NA	20.1	10.03	01/28/2021	11:02
4	ICV	1	Water	NA	NA	20.2	7.00	01/28/2021	11:05
5	CCV1	1	Water	NA	NA	20.2	2.01	01/28/2021	11:06
6	M1266-01	1	Water	NA	NA	20.1	5.42	01/28/2021	11:10
7	M1266-01DUP	1	Water	NA	NA	20.1	5.43	01/28/2021	11:11
8	CCV2	1	Water	NA	NA	20.1	12.02	01/28/2021	11:12

WORKLIST(Hardcopy Internal Chain)

2B 112951

WorkList Name : PH M1266 **WorkList ID :** 146036 **Department :** Wet-Chemistry **Date :** 01-28-2021 10:30:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/27/2021	Water	M1266-01-1	pH	Cool 4 deg C	STER05	G32	BOTTLE-1	01/27/2021	SM 4500-H B

Date/Time 01/26/2021 10:15:55
Received by: [Signature]
Relinquished by: [Signature]

Date/Time 01/26/2021 11:10
Received by: [Signature]
Relinquished by: [Signature]