

LB 112952


7471B

INSTRUMENT ID: CV1

Linear

 $\mu$  Abs.:

59133



A= 0.0000e+000

B= 1.6904e-004

C= 5.4427e-003

Rho= 0.9999960

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|----|
| 0.0    | 0.000  | 0.009  | 0.009  | 22    | 0.000      | 22    |       |       |       |       |    |
| 0.2    | 0.200  | 0.182  | -0.018 | 1045  | 0.0 %      | 1045  |       |       |       |       | -9 |
| 2.5    | 2.500  | 2.515  | 0.015  | 14845 | 0.0 %      | 14845 |       |       |       |       | 1  |
| 5.0    | 5.000  | 4.997  | -0.003 | 29527 | 0.0 %      | 29527 |       |       |       |       | 0  |
| 7.5    | 7.500  | 7.496  | -0.004 | 44315 | 0.0 %      | 44315 |       |       |       |       | 0  |
| 10.0   | 10.000 | 10.001 | 0.001  | 59133 | 0.0 %      | 59133 |       |       |       |       | 0  |

## LB112952 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 28 Jan 2021 11:57:57

| Sample ID        | Extended ID    | $\mu$ Abs. | Conc.   | Std Conc     | Method | Units | Date                 | Type | Type |
|------------------|----------------|------------|---------|--------------|--------|-------|----------------------|------|------|
| 0.0 - 1          | 50             | 22         | -       | 0.00007471B  | PPB    |       | 28 Jan 2021 12:01:22 | S    | Std  |
| 0.2 - 1          | 50.2           | 1045       | -       | 0.20007471B  | PPB    |       | 28 Jan 2021 12:03:39 | S    | Std  |
| 2.5 - 1          | 52.5           | 14845      | -       | 2.50007471B  | PPB    |       | 28 Jan 2021 12:05:56 | S    | Std  |
| 5.0 - 1          | 55             | 29527      | -       | 5.00007471B  | PPB    |       | 28 Jan 2021 12:08:13 | S    | Std  |
| 7.5 - 1          | 57.5           | 44315      | -       | 7.50007471B  | PPB    |       | 28 Jan 2021 12:10:30 | S    | Std  |
| 10.0 - 1         | 510            | 59133      | -       | 10.00007471B | PPB    |       | 28 Jan 2021 12:12:45 | S    | Std  |
| ICV86 - 1        | ICV86          | 23067      | 3.9046  | -7471B       | PPB    |       | 28 Jan 2021 12:15:30 | U    | SMPL |
| ICB86 - 1        | ICB86          | -54        | -0.0037 | -7471B       | PPB    |       | 28 Jan 2021 12:17:45 | U    | SMPL |
| CCV01 - 1        | CCV01          | 29557      | 5.0017  | -7471B       | PPB    |       | 28 Jan 2021 12:20:00 | U    | SMPL |
| CCB01 - 1        | CCB01          | -55        | -0.0039 | -7471B       | PPB    |       | 28 Jan 2021 12:22:15 | U    | SMPL |
| CRA - 1          | CRA            | 1047       | 0.1824  | -7471B       | PPB    |       | 28 Jan 2021 12:24:31 | U    | SMPL |
| HighStd - 1      | HighStd        | 59361      | 10.0396 | -7471B       | PPB    |       | 28 Jan 2021 12:26:46 | U    | SMPL |
| ChkStd - 1       | ChkStd         | 40594      | 6.8673  | -7471B       | PPB    |       | 28 Jan 2021 12:31:23 | U    | SMPL |
| PB134380BL - 1   | PBS            | -68        | -0.0061 | -7471B       | PPB    |       | 28 Jan 2021 12:33:38 | U    | SMPL |
| PB134380BS - 1   | LCSS           | 21644      | 3.6641  | -7471B       | PPB    |       | 28 Jan 2021 12:35:54 | U    | SMPL |
| M1250-01 - 1     | T1-CONTAINMENT | 789        | 0.1388  | -7471B       | PPB    |       | 28 Jan 2021 12:38:11 | U    | SMPL |
| M1251-01 - 1     | 400026         | 1690       | 0.2911  | -7471B       | PPB    |       | 28 Jan 2021 12:40:28 | U    | SMPL |
| M1251-03 - 1     | 200010         | 2546       | 0.4358  | -7471B       | PPB    |       | 28 Jan 2021 12:42:45 | U    | SMPL |
| M1251-05 - 1     | 200059         | 3352       | 0.5721  | -7471B       | PPB    |       | 28 Jan 2021 12:45:01 | U    | SMPL |
| M1251-07 - 1     | 400024         | 2479       | 0.4245  | -7471B       | PPB    |       | 28 Jan 2021 12:47:17 | U    | SMPL |
| CCV02 - 1        | CCV02          | 29269      | 4.9530  | -7471B       | PPB    |       | 28 Jan 2021 12:49:32 | U    | SMPL |
| CCB02 - 1        | CCB02          | -37        | -0.0008 | -7471B       | PPB    |       | 28 Jan 2021 12:51:47 | U    | SMPL |
| M1251-07DUP - 1  | 400024DUP      | 2230       | 0.3824  | -7471B       | PPB    |       | 28 Jan 2021 12:54:03 | U    | SMPL |
| M1251-07MS - 1   | 400024MS       | 22278      | 3.7712  | -7471B       | PPB    |       | 28 Jan 2021 12:56:18 | U    | SMPL |
| M1251-07MSD - 1  | 400024MSD      | 23404      | 3.9616  | -7471B       | PPB    |       | 28 Jan 2021 12:58:34 | U    | SMPL |
| M1252-01 - 1     | BS-SP          | 2331       | 0.3995  | -7471B       | PPB    |       | 28 Jan 2021 13:00:50 | U    | SMPL |
| M1253-01 - 1     | 124055         | 1419       | 0.2453  | -7471B       | PPB    |       | 28 Jan 2021 13:03:07 | U    | SMPL |
| M1253-03 - 1     | 124035         | 802        | 0.1410  | -7471B       | PPB    |       | 28 Jan 2021 13:05:24 | U    | SMPL |
| M1253-05 - 1     | VNJ-201        | 1731       | 0.2980  | -7471B       | PPB    |       | 28 Jan 2021 13:07:40 | U    | SMPL |
| M1253-07 - 1     | VNJ-229        | 3567       | 0.6084  | -7471B       | PPB    |       | 28 Jan 2021 13:09:57 | U    | SMPL |
| M1254-01 - 1     | CRT-15         | 104120     | 17.6055 | -7471B       | PPB    |       | 28 Jan 2021 13:12:15 | U    | SMPL |
| M1254-06 - 1     | CRT-5          | 1518       | 0.2620  | -7471B       | PPB    |       | 28 Jan 2021 13:14:30 | U    | SMPL |
| CCV03 - 1        | CCV03          | 29362      | 4.9687  | -7471B       | PPB    |       | 28 Jan 2021 13:16:48 | U    | SMPL |
| CCB03 - 1        | CCB03          | -49        | -0.0028 | -7471B       | PPB    |       | 28 Jan 2021 13:19:04 | U    | SMPL |
| M1255-01 - 1     | CRT-3          | 4475       | 0.7619  | -7471B       | PPB    |       | 28 Jan 2021 13:21:20 | U    | SMPL |
| M1255-06 - 1     | CRT-13         | 18077      | 3.0611  | -7471B       | PPB    |       | 28 Jan 2021 13:23:35 | U    | SMPL |
| M1255-11 - 1     | CRT-17         | 1425       | 0.2463  | -7471B       | PPB    |       | 28 Jan 2021 13:25:51 | U    | SMPL |
| M1256-02 - 1     | SIDEWALK       | 26469      | 4.4797  | -7471B       | PPB    |       | 28 Jan 2021 13:28:06 | U    | SMPL |
| M1251-07LX5 - 1  |                | 429        | 0.0780  | -7471B       | PPB    |       | 28 Jan 2021 13:30:22 | U    | SMPL |
| M1251-07A - 1    |                | 27381      | 4.6338  | -7471B       | PPB    |       | 28 Jan 2021 13:32:38 | U    | SMPL |
| M1254-01DLX5 - 1 |                | 20343      | 3.4442  | -7471B       | PPB    |       | 28 Jan 2021 13:34:55 | U    | SMPL |
| CCV04 - 1        | CCV04          | 29465      | 4.9861  | -7471B       | PPB    |       | 28 Jan 2021 13:37:12 | U    | SMPL |
| CCB04 - 1        | CCB04          | -61        | -0.0049 | -7471B       | PPB    |       | 28 Jan 2021 13:39:29 | U    | SMPL |