

LB112960


SFAM01.1

INSTRUMENT ID: CV1

Linear

 $\mu$  Abs.:

48668



A= 0.0000e+000

B= 2.0527e-004 *slope*C= -1.4975e-002 *y-intercept*

Rho= 0.9999788

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D  |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|-----|
| 0.00   | 0.000  | -0.002 | -0.002 | 62    | 0.000      | 62    |       |       |       |       | =   |
| 0.05   | 0.050  |        |        |       |            | 334   |       |       |       |       | -13 |
| 0.20   | 0.200  | 0.174  | -0.026 | 919   | 0.0 %      | 919   |       |       |       |       | 0   |
| 2.50   | 2.500  | 2.507  | 0.007  | 12287 | 0.0 %      | 12287 |       |       |       |       | 0   |
| 5.00   | 5.000  | 5.046  | 0.046  | 24654 | 0.0 %      | 24654 |       |       |       |       | 1   |
| 7.50   | 7.500  | 7.500  | 0.000  | 36610 | 0.0 %      | 36610 |       |       |       |       | 0   |
| 10.0   | 10.000 | 9.975  | -0.025 | 48668 | 0.0 %      | 48668 |       |       |       |       | 0   |

LB112960  
INSTRUMENT ID : CV1

| Sample ID  | Extended ID | $\mu$ Abs. | Conc.     | Std Conc | Method       | Units | Date            | Type |
|------------|-------------|------------|-----------|----------|--------------|-------|-----------------|------|
|            | 0 S0        | 62         | -         |          | 0 SFAM01.1   | PPB   | 1/28/2021 15:19 | Std  |
|            | 0.2 S01     | 919        | -         |          | 0.2 SFAM01.1 | PPB   | 1/28/2021 15:21 | Std  |
|            | 2.5 S02     | 12287      | -         |          | 2.5 SFAM01.1 | PPB   | 1/28/2021 15:23 | Std  |
|            | 5 S03       | 24654      | -         |          | 5 SFAM01.1   | PPB   | 1/28/2021 15:26 | Std  |
|            | 7.5 S04     | 36610      | -         |          | 7.5 SFAM01.1 | PPB   | 1/28/2021 15:28 | Std  |
|            | 10 S05      | 48668      | -         |          | 10 SFAM01.1  | PPB   | 1/28/2021 15:30 | Std  |
| ICV012     | ICV012      | 18724      | 3.8286 -  |          | SFAM01.1     | PPB   | 1/28/2021 15:33 | SMPL |
| ICB012     | ICB012      | -21        | -0.0193 - |          | SFAM01.1     | PPB   | 1/28/2021 15:35 | SMPL |
| CCV005     | CCV005      | 24897      | 5.0958 -  |          | SFAM01.1     | PPB   | 1/28/2021 15:38 | SMPL |
| CCB005     | CCB005      | -27        | -0.0205 - |          | SFAM01.1     | PPB   | 1/28/2021 15:40 | SMPL |
| PB134381BL | PBS381      | -1         | -0.0152 - |          | SFAM01.1     | PPB   | 1/28/2021 15:42 | SMPL |
| M1204-19   | X2X99       | 35720      | 7.3174 -  |          | SFAM01.1     | PPB   | 1/28/2021 15:47 | SMPL |
| M1204-20   | X2X99D      | 35663      | 7.3057 -  |          | SFAM01.1     | PPB   | 1/28/2021 15:49 | SMPL |
| M1204-21   | X2X99S      | 47591      | 9.7543 -  |          | SFAM01.1     | PPB   | 1/28/2021 15:56 | SMPL |
| CCV006     | CCV006      | 24381      | 4.9898 -  |          | SFAM01.1     | PPB   | 1/28/2021 15:58 | SMPL |
| CCB006     | CCB006      | -22        | -0.0195 - |          | SFAM01.1     | PPB   | 1/28/2021 16:00 | SMPL |