



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: jignesh

Date: 01/29/2021

OVEN IN TEMP1: 103 °C

OVEN OUT TEMP1: 104 °C

DATE & TIME IN1: 01/28/2021 16:30 DATE & TIME OUT1: 01/29/2021 08:20

OVEN IN TEMP2: 104 °C

OVEN OUT TEMP2: 104 °C

DATE & TIME IN2: 01/29/2021 10:00 DATE & TIME OUT2: 01/29/2021 10:35

Run Number: LB112962

BalanceID: WC SC-3

OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB112962BL	LB112962BL	87.4139	87.4139	100	87.4139	87.4139	87.4139	0.0000	0
2	M1278-01	TRANSIT-AUTHORITY-149TH-GRAN	85.0829	85.0829	100	85.1781	85.1781	85.1781	0.0952	952

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

WORKLIST(Hardcopy Internal Chain)

18112962

WorkList Name : ts m1268 WorkList ID : 146048 Department : Wet-Chemistry Date : 01-28-2021 16:04:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/04/2021	Water	M1278-01	TS	Cool 4 deg C	TULL02		TRANSIT-AUTHORITY-149T	01/28/2021	SM2540 B

Date/Time 01.28.2021 16:10
 Received by: JD/est (em)
 Relinquished by: KS (wetchen)

Date/Time 01.28.2021 17:20
 Received by: KS (wetchen)
 Relinquished by: JD/est

Chen