


LB 112979

7470A

INSTRUMENT ID: CV1

 μ Abs.:

40251



Linear

A= 0.0000e+000

B= 2.4940e-004

C= 1.3344e-002

Rho= 0.9999545

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/1/1
0.0	0.000	0.020	0.020	27	0.000	27					1
0.2	0.200	0.204	0.004	765	0.0 %	765					2
2.5	2.500	2.494	-0.006	9945	0.0 %	9945					0
5.0	5.000	4.995	-0.005	19976	0.0 %	19976					0
7.5	7.500	7.435	-0.065	29757	0.0 %	29757					1
10.0	10.000	10.052	0.052	40251	0.0 %	40251					1

LB112979 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 29 Jan 2021 10:39:42

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	27	-	0.0000	7470A	PPB	29 Jan 2021 10:41:58	S	Std
0.2 - 1	50.2	765	-	0.2000	7470A	PPB	29 Jan 2021 10:44:16	S	Std
2.5 - 1	52.5	9945	-	2.5000	7470A	PPB	29 Jan 2021 10:46:33	S	Std
5.0 - 1	55	19976	-	5.0000	7470A	PPB	29 Jan 2021 10:48:50	S	Std
7.5 - 1	57.5	29757	-	7.5000	7470A	PPB	29 Jan 2021 10:51:06	S	Std
10.0 - 1	57.5	40251	-	10.0000	7470A	PPB	29 Jan 2021 10:55:56	S	Std
ICV87 - 1	ICV87	15172	3.7973	-	7470A	PPB	29 Jan 2021 10:58:32	U	SMPL
ICB87 - 1	ICB87	-93	-0.0099	-	7470A	PPB	29 Jan 2021 11:00:47	U	SMPL
CCV07 - 1	CCV07	20532	5.1340	-	7470A	PPB	29 Jan 2021 11:03:02	U	SMPL
CCB07 - 1	CCB07	-85	-0.0079	-	7470A	PPB	29 Jan 2021 11:05:17	U	SMPL
CRA - 1	CRA	728	0.1949	-	7470A	PPB	29 Jan 2021 11:07:32	U	SMPL
HighStd - 1	HighStd	40481	10.1094	-	7470A	PPB	29 Jan 2021 11:09:48	U	SMPL
ChkStd - 1	ChkStd	27834	6.9552	-	7470A	PPB	29 Jan 2021 11:12:03	U	SMPL
PB134390BL - 1	PBW	-114	-0.0151	-	7470A	PPB	29 Jan 2021 11:14:20	U	SMPL
PB134390BS - 1	LCSW	15203	3.8050	-	7470A	PPB	29 Jan 2021 11:16:36	U	SMPL
M1174-02 - 1	OWS-1	-32	0.0054	-	7470A	PPB	29 Jan 2021 11:18:53	U	SMPL
M1241-02 - 1	FACILITY-1	0	0.0133	-	7470A	PPB	29 Jan 2021 11:21:09	U	SMPL
M1254-02 - 1	CRT-15	83	0.0340	-	7470A	PPB	29 Jan 2021 11:23:27	U	SMPL
M1254-07 - 1	CRT-5	20	0.0183	-	7470A	PPB	29 Jan 2021 11:25:44	U	SMPL
M1255-02 - 1	CRT-3	59	0.0281	-	7470A	PPB	29 Jan 2021 11:27:59	U	SMPL
CCV08 - 1	CCV08	20396	5.1001	-	7470A	PPB	29 Jan 2021 11:30:14	U	SMPL
CCB08 - 1	CCB08	-63	-0.0024	-	7470A	PPB	29 Jan 2021 11:32:30	U	SMPL
M1255-07 - 1	CRT-13	0	0.0133	-	7470A	PPB	29 Jan 2021 11:34:45	U	SMPL
M1255-12 - 1	CRT-17	18	0.0178	-	7470A	PPB	29 Jan 2021 11:37:01	U	SMPL
M1262-01 - 1	JTY2-WC-S-1-1.0-6.0	31	0.0211	-	7470A	PPB	29 Jan 2021 11:39:16	U	SMPL
M1262-01DUP - 1	JTY2-WC-S-1-1.0-6.0DUP	2	0.0138	-	7470A	PPB	29 Jan 2021 11:41:32	U	SMPL
M1262-01MS - 1	JTY2-WC-S-1-1.0-6.0MS	15116	3.7833	-	7470A	PPB	29 Jan 2021 11:43:48	U	SMPL
M1262-01MSD - 1	JTY2-WC-S-1-1.0-6.0MSD	15139	3.7890	-	7470A	PPB	29 Jan 2021 11:46:04	U	SMPL
PB134391BL - 1	PBW	59	0.0281	-	7470A	PPB	29 Jan 2021 11:48:21	U	SMPL
PB134391BS - 1	LCSW	15953	3.9920	-	7470A	PPB	29 Jan 2021 11:50:38	U	SMPL
M1242-02 - 1	0026-SITE-WATER-048-03-04	141	0.0485	-	7470A	PPB	29 Jan 2021 11:52:55	U	SMPL
M1263-01 - 1	0048-042-COMP-B-ELUTRIATE	80	0.0333	-	7470A	PPB	29 Jan 2021 11:55:10	U	SMPL
CCV09 - 1	CCV09	20622	5.1565	-	7470A	PPB	29 Jan 2021 11:57:26	U	SMPL
CCB09 - 1	CCB09	-67	-0.0034	-	7470A	PPB	29 Jan 2021 11:59:41	U	SMPL
M1263-01DUP - 1	0048-042-COMP-B-ELUTRIATEDUP	71	0.0311	-	7470A	PPB	29 Jan 2021 12:01:57	U	SMPL
M1263-01MS - 1	0048-042-COMP-B-ELUTRIATEMS	17852	4.4657	-	7470A	PPB	29 Jan 2021 12:04:13	U	SMPL
M1263-01MSD - 1	0048-042-COMP-B-ELUTRIATEMSD	15940	3.9888	-	7470A	PPB	29 Jan 2021 12:06:28	U	SMPL
M1263-02 - 1	0053-048-COMP-G-ELUTRIATE	28	0.0203	-	7470A	PPB	29 Jan 2021 12:08:44	U	SMPL
M1263-03 - 1	0054-051-COMP-H-ELUTRIATE	44	0.0243	-	7470A	PPB	29 Jan 2021 12:11:00	U	SMPL
M1263-04 - 1	0048-042-COMP-B-ELUTRIATE	44	0.0243	-	7470A	PPB	29 Jan 2021 12:13:17	U	SMPL
M1263-05 - 1	0053-048-COMP-G-ELUTRIATE	42	0.0238	-	7470A	PPB	29 Jan 2021 12:15:33	U	SMPL
M1263-06 - 1	0054-051-COMP-H-ELUTRIATE	57	0.0276	-	7470A	PPB	29 Jan 2021 12:17:50	U	SMPL
PB134390TB - 1	PB134371TB	-11	0.0106	-	7470A	PPB	29 Jan 2021 12:20:07	U	SMPL
CCV10 - 1	CCV10	20495	5.1248	-	7470A	PPB	29 Jan 2021 12:22:22	U	SMPL
CCB10 - 1	CCB10	-67	-0.0034	-	7470A	PPB	29 Jan 2021 12:24:38	U	SMPL
M1262-01LX5 - 1		-1	0.0131	-	7470A	PPB	29 Jan 2021 12:26:54	U	SMPL
M1262-01A - 1		15935	3.9876	-	7470A	PPB	29 Jan 2021 12:29:09	U	SMPL
M1263-01LX5 - 1		-23	0.0076	-	7470A	PPB	29 Jan 2021 12:31:25	U	SMPL
M1263-01A - 1		16968	4.2452	-	7470A	PPB	29 Jan 2021 12:33:41	U	SMPL
CCV11 - 1	CCV11	20417	5.1054	-	7470A	PPB	29 Jan 2021 12:35:57	U	SMPL
CCB11 - 1	CCB11	-71	-0.0044	-	7470A	PPB	29 Jan 2021 12:38:12	U	SMPL