



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112982

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/29/2021	13:15
2	CAL2	1	Water	NA	NA	20.2	7.00	01/29/2021	13:16
3	CAL3	1	Water	NA	NA	20.2	10.03	01/29/2021	13:17
4	ICV	1	Water	NA	NA	20.2	7.00	01/29/2021	13:19
5	CCV1	1	Water	NA	NA	20.2	2.01	01/29/2021	13:20
6	M1293-01	1	Solid	20.03	20	20.2	8.97	01/29/2021	13:22
7	M1293-01DUP	1	Solid	20.04	20	20.3	8.98	01/29/2021	13:23
8	CCV2	1	Water	NA	NA	20.2	12.02	01/29/2021	13:25

WORKLIST(Hardcopy Internal Chain)

12982

WorkList Name : PH M1293 WorkList ID : 146084 Department : Wet-Chemistry Date : 01-29-2021 13:08:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/29/2021	Solid	M1293-01	pH	Cool 4 deg C	TULL02	G41	TEST-PIT-4	01/29/2021	9045D

Date/Time 01.29.2021 13:10
 Received by: To wet Chem
 Relinquished by: KS (wetchem)

Date/Time 01.29.2021 14:30
 Received by: KS (wetchem)
 Relinquished by: To wet Chem