


LB112985

7471B

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.7117e-004

C= 2.7732e-002

Rho= 0.9999463

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.030	0.030	15	0.000	15					1
0.2	0.200	0.199	-0.001	998	0.0 %	998					1
2.5	2.500	2.482	-0.018	14337	0.0 %	14337					1
5.0	5.000	5.006	0.006	29082	0.0 %	29082					0
7.5	7.500	7.431	-0.069	43252	0.0 %	43252					1
10.0	10.000	10.052	0.052	58566	0.0 %	58566					1

LB112985 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 29 Jan 2021 14:25:07

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	15	-	0.00007471B	PPB		29 Jan 2021 14:26:25	S	Std
0.2 - 1	50.2	998	-	0.20007471B	PPB		29 Jan 2021 14:28:42	S	Std
2.5 - 1	52.5	14337	-	2.50007471B	PPB		29 Jan 2021 14:30:59	S	Std
5.0 - 1	55	29082	-	5.00007471B	PPB		29 Jan 2021 14:33:16	S	Std
7.5 - 1	57.5	43252	-	7.50007471B	PPB		29 Jan 2021 14:35:33	S	Std
10.0 - 1	510	58566	-	10.00007471B	PPB		29 Jan 2021 14:37:49	S	Std
ICV88 - 1	ICV88	21392	3.6894	-7471B	PPB		29 Jan 2021 14:40:32	U	SMPL
ICB88 - 1	ICB88	-67	0.0163	-7471B	PPB		29 Jan 2021 14:42:46	U	SMPL
CCV12 - 1	CCV12	28953	4.9836	-7471B	PPB		29 Jan 2021 14:45:02	U	SMPL
CCB12 - 1	CCB12	-61	0.0173	-7471B	PPB		29 Jan 2021 14:47:17	U	SMPL
CRA - 1	CRA	966	0.1931	-7471B	PPB		29 Jan 2021 14:49:32	U	SMPL
HighStd - 1	HighStd	59180	10.1576	-7471B	PPB		29 Jan 2021 14:51:48	U	SMPL
ChkStd - 1	ChkStd	42044	7.2244	-7471B	PPB		29 Jan 2021 14:54:03	U	SMPL
PB134421BL - 1	PBS	-126	0.0062	-7471B	PPB		29 Jan 2021 14:56:20	U	SMPL
PB134421BS - 1	LCSS	21101	3.6396	-7471B	PPB		29 Jan 2021 14:58:37	U	SMPL
M1267-02 - 1	1-3-DRUMS	254	0.0712	-7471B	PPB		29 Jan 2021 15:00:54	U	SMPL
M1269-01 - 1	CL-02-012821	4421	0.7845	-7471B	PPB		29 Jan 2021 15:03:11	U	SMPL
M1270-01 - 1	RS-SP1	1968	0.3646	-7471B	PPB		29 Jan 2021 15:05:27	U	SMPL
M1270-03 - 1	RS-SP2	2171	0.3993	-7471B	PPB		29 Jan 2021 15:07:44	U	SMPL
M1270-05 - 1	RS-SP3	1334	0.2561	-7471B	PPB		29 Jan 2021 15:09:59	U	SMPL
CCV13 - 1	CCV13	28928	4.9793	-7471B	PPB		29 Jan 2021 15:12:15	U	SMPL
CCB13 - 1	CCB13	-52	0.0188	-7471B	PPB		29 Jan 2021 15:14:30	U	SMPL
M1271-01 - 1	RO-332	3040	0.5481	-7471B	PPB		29 Jan 2021 15:16:46	U	SMPL
M1271-03 - 1	RO-RBR-200035	7530	1.3166	-7471B	PPB		29 Jan 2021 15:19:01	U	SMPL
M1271-03DUP - 1	RO-RBR-200035DUP	8109	1.4157	-7471B	PPB		29 Jan 2021 15:21:16	U	SMPL
M1271-03MS - 1	RO-RBR-200035MS	29795	5.1277	-7471B	PPB		29 Jan 2021 15:23:32	U	SMPL
M1271-03MSD - 1	RO-RBR-200035MSD	30878	5.3131	-7471B	PPB		29 Jan 2021 15:25:48	U	SMPL
M1271-03LX5 - 1		1401	0.2675	-7471B	PPB		29 Jan 2021 15:28:05	U	SMPL
M1271-03A - 1		30547	5.2565	-7471B	PPB		29 Jan 2021 15:30:22	U	SMPL
CCV14 - 1	CCV14	29479	5.0737	-7471B	PPB		29 Jan 2021 15:32:39	U	SMPL
CCB14 - 1	CCB14	-59	0.0176	-7471B	PPB		29 Jan 2021 15:34:56	U	SMPL