


LB 113005

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.012	-0.012	-2	0.000	-2					14
0.2	0.200	0.173	-0.027	1020	0.0 %	1020					1
2.5	2.500	2.482	-0.018	13817	0.0 %	13817					3
5.0	5.000	5.174	0.174	28729	0.0 %	28729					2
7.5	7.500	7.377	-0.123	40935	0.0 %	40935					0
10.0	10.000	10.006	0.006	55505	0.0 %	55505					

LB113005 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 Feb 2021 10:06:27

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-2	-	0.0000	7471B	PPB	03 Feb 2021 10:30:11	S	Std
0.2 - 1	50.2	1020	-	0.2000	7471B	PPB	03 Feb 2021 10:37:52	S	Std
2.5 - 1	52.5	13817	-	2.5000	7471B	PPB	03 Feb 2021 10:40:09	S	Std
5.0 - 1	55	28729	-	5.0000	7471B	PPB	03 Feb 2021 10:42:26	S	Std
7.5 - 1	57.5	40935	-	7.5000	7471B	PPB	03 Feb 2021 10:44:43	S	Std
10.0 - 1	59	55505	-	10.0000	7471B	PPB	03 Feb 2021 10:49:37	S	Std
ICV89 - 1	ICV89	22837	4.1103	-	7471B	PPB	03 Feb 2021 10:53:28	U	SMPL
ICB89 - 1	ICB89	-93	-0.0280	-	7471B	PPB	03 Feb 2021 10:55:43	U	SMPL
CCV15 - 1	CCV15	28805	5.1874	-	7471B	PPB	03 Feb 2021 10:57:58	U	SMPL
CCB15 - 1	CCB15	-109	-0.0309	-	7471B	PPB	03 Feb 2021 11:00:13	U	SMPL
CRA - 1	CRA	999	0.1691	-	7471B	PPB	03 Feb 2021 11:02:29	U	SMPL
HighStd - 1	HighStd	55989	10.0934	-	7471B	PPB	03 Feb 2021 11:04:44	U	SMPL
ChkStd - 1	ChkStd	39666	7.1475	-	7471B	PPB	03 Feb 2021 11:07:00	U	SMPL
PB134427BL - 1	PBS	-75	-0.0247	-	7471B	PPB	03 Feb 2021 11:09:16	U	SMPL
PB134427BS - 1	LCSS	21645	3.8952	-	7471B	PPB	03 Feb 2021 11:11:34	U	SMPL
M1284-01 - 1	TR-04-012921	1378	0.2375	-	7471B	PPB	03 Feb 2021 11:13:50	U	SMPL
M1286-01 - 1	JJ_BH_01_0-1	14645	2.6319	-	7471B	PPB	03 Feb 2021 11:16:07	U	SMPL
M1286-01DUP - 1	JJ_BH_01_0-1DUP	14976	2.6916	-	7471B	PPB	03 Feb 2021 11:21:06	U	SMPL
M1286-01MS - 1	JJ_BH_01_0-1MS	38683	6.9701	-	7471B	PPB	03 Feb 2021 11:23:23	U	SMPL
M1286-01MSD - 1	JJ_BH_01_0-1MSD	37091	6.6828	-	7471B	PPB	03 Feb 2021 11:25:38	U	SMPL
CCV16 - 1	CCV16	29106	5.2417	-	7471B	PPB	03 Feb 2021 11:27:55	U	SMPL
CCB16 - 1	CCB16	-89	-0.0273	-	7471B	PPB	03 Feb 2021 11:30:11	U	SMPL
M1286-03 - 1	JJ_BH_01_2-4	2841	0.5015	-	7471B	PPB	03 Feb 2021 11:32:27	U	SMPL
M1286-05 - 1	JJ_BH_01_4-6	3538	0.6273	-	7471B	PPB	03 Feb 2021 11:34:44	U	SMPL
M1286-07 - 1	JJ_BH_01_6-7	1280	0.2198	-	7471B	PPB	03 Feb 2021 11:36:59	U	SMPL
M1286-09 - 1	JJ_BH_02_0-1	3029	0.5355	-	7471B	PPB	03 Feb 2021 11:39:15	U	SMPL
M1286-11 - 1	JJ_BH_02_2-4	1591	0.2759	-	7471B	PPB	03 Feb 2021 11:41:32	U	SMPL
M1286-13 - 1	JJ_BH_02_4-6	9462	1.6965	-	7471B	PPB	03 Feb 2021 11:43:48	U	SMPL
M1286-15 - 1	JJ_BH_02_6-7	1022	0.1733	-	7471B	PPB	03 Feb 2021 11:46:04	U	SMPL
M1286-17 - 1	JJ_BH_03_0-1	2981	0.5268	-	7471B	PPB	03 Feb 2021 11:48:22	U	SMPL
M1286-19 - 1	JJ_BH_03_2-4	3739	0.6636	-	7471B	PPB	03 Feb 2021 11:50:39	U	SMPL
M1286-21 - 1	JJ_BH_03_4-6	1276	0.2191	-	7471B	PPB	03 Feb 2021 11:52:55	U	SMPL
CCV17 - 1	CCV17	28752	5.1778	-	7471B	PPB	03 Feb 2021 11:55:11	U	SMPL
CCB17 - 1	CCB17	-73	-0.0244	-	7471B	PPB	03 Feb 2021 11:57:27	U	SMPL
M1286-23 - 1	JJ_BH_03_6-7	3265	0.5781	-	7471B	PPB	03 Feb 2021 11:59:43	U	SMPL
M1286-01LX5 - 1		2911	0.5142	-	7471B	PPB	03 Feb 2021 12:01:58	U	SMPL
M1286-01A - 1		37483	6.7535	-	7471B	PPB	03 Feb 2021 12:04:14	U	SMPL
CCV18 - 1	CCV18	28900	5.2045	-	7471B	PPB	03 Feb 2021 12:06:29	U	SMPL
CCB18 - 1	CCB18	-80	-0.0256	-	7471B	PPB	03 Feb 2021 12:08:47	U	SMPL