


LB 113008

245.1

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.021	-0.021	29	0.000	29					
0.05	0.050	0.051	0.001	290	0.0 %	290	0				
0.2	0.200	0.200	-0.000	833	0.0 %	833					
2.5	2.500	2.528	0.028	9323	0.0 %	9323					
5.0	5.000	4.987	-0.013	18291	0.0 %	18291					
7.5	7.500	7.523	0.023	27538	0.0 %	27538					
10.0	10.000	9.982	-0.018	36504	0.0 %	36504					

100-021

LB113008 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 03 Feb 2021 13:47:17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	29	-	0.0000	245.1	PPB	03 Feb 2021 13:56:00	S	Std
0.05 - 1	50.05	290	-	0.0500	245.1	PPB	03 Feb 2021 13:58:17	S	Std
0.2 - 1	50.2	833	-	0.2000	245.1	PPB	03 Feb 2021 14:00:35	S	Std
2.5 - 1	52.5	9323	-	2.5000	245.1	PPB	03 Feb 2021 14:02:52	S	Std
5.0 - 1	55	18291	-	5.0000	245.1	PPB	03 Feb 2021 14:05:09	S	Std
7.5 - 1	57.5	27538	-	7.5000	245.1	PPB	03 Feb 2021 14:07:24	S	Std
10.0 - 1	510	36504	-	10.0000	245.1	PPB	03 Feb 2021 14:14:53	S	Std
ICV90 - 1	ICV90	14014	3.8144	-245.1		PPB	03 Feb 2021 14:17:39	U	SMPL
ICB90 - 1	ICB90	56	-0.0133	-245.1		PPB	03 Feb 2021 14:19:54	U	SMPL
CCV19 - 1	CCV19	18712	5.1027	-245.1		PPB	03 Feb 2021 14:22:10	U	SMPL
CCB19 - 1	CCB19	51	-0.0147	-245.1		PPB	03 Feb 2021 14:24:25	U	SMPL
CRA - 1	CRA	262	0.0432	-245.1		PPB	03 Feb 2021 14:26:41	U	SMPL
HighStd - 1	HighStd	36474	9.9736	-245.1		PPB	03 Feb 2021 14:28:57	U	SMPL
ChkStd - 1	ChkStd	26231	7.1647	-245.1		PPB	03 Feb 2021 14:31:12	U	SMPL
PB134449BL - 1	PBW	49	-0.0153	-245.1		PPB	03 Feb 2021 14:36:09	U	SMPL
PB134449BS - 1	LCSW	14438	3.9307	-245.1		PPB	03 Feb 2021 14:38:25	U	SMPL
M1278-01 - 1	149TH-ST-GW-1	962	0.2351	-245.1		PPB	03 Feb 2021 14:40:41	U	SMPL
M1278-01DUP - 1	149TH-ST-GW-1DUP	1033	0.2546	-245.1		PPB	03 Feb 2021 14:45:31	U	SMPL
M1278-01MS - 1	149TH-ST-GW-1MS	15844	4.3162	-245.1		PPB	03 Feb 2021 14:50:41	U	SMPL
M1278-01MSD - 1	149TH-ST-GW-1MSD	16035	4.3686	-245.1		PPB	03 Feb 2021 14:52:58	U	SMPL
CCV20 - 1	CCV20	18155	4.9500	-245.1		PPB	03 Feb 2021 14:55:13	U	SMPL
CCB20 - 1	CCB20	39	-0.0180	-245.1		PPB	03 Feb 2021 14:57:29	U	SMPL
M1278-01LX5 - 1		248	0.0393	-245.1		PPB	03 Feb 2021 14:59:45	U	SMPL
M1278-01A - 1		16018	4.3639	-245.1		PPB	03 Feb 2021 15:02:00	U	SMPL
CCV21 - 1	CCV21	18387	5.0136	-245.1		PPB	03 Feb 2021 15:04:15	U	SMPL
CCB21 - 1	CCB21	36	-0.0188	-245.1		PPB	03 Feb 2021 15:06:31	U	SMPL