



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB113013

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: FLASHPONT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	83.0	02/03/2021	09:35
2	M1278-01		1	9	100.00	>212.0	>212.0	02/03/2021	10:10
3	M1278-01DUP		1	9	100.00	>212.0	>212.0	02/03/2021	10:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB113013

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP12721 WorkList ID : 146006 Department : Wet-Chemistry Date : 01-27-2021 11:31:13

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/04/2021	Water	M1278-01	Flash Point	Cool 4 deg C	TULL02		149TH-ST-GW-1	01/28/2021	1010A

Date/Time 02/03/21 09:05
Received by: KS (Wetchem)
Relinquished by: KS (Wetchem)

Date/Time 02/03/21 10:45
Received by: KS (Wetchem)
Relinquished by: KS (Wetchem)