



PERCENT SOLID

Supervisor: apatel
Analyst: ketankumar
Date: 2/4/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 16:50
In Date: 02/03/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 108
Time OUT: 09:15
Out Date: 02/04/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB113015

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
M1300-01	MGB2N0	1	1.17	9.75	9.51	97.2	
M1300-02	MGB2N1	2	1.16	9.51	9.25	96.9	
M1300-03	MGB2N2	3	1.16	9.9	9.51	95.5	
M1300-04	MGB2N3	4	1.16	9.51	9.06	94.6	
M1300-05	MGB2N4	5	1.16	9.78	9.58	97.7	
M1300-06	MGB2N4D	6	1.16	9.78	9.58	97.7	
M1300-07	MGB2N4S	7	1.16	9.78	9.58	97.7	
M1300-08	MGB2N5	8	1.16	9.92	9.72	97.7	
M1300-09	MGB2N6	9	1.17	9.78	9.56	97.4	
M1300-10	MGB2N7	10	1.16	9.65	9.41	97.2	
M1300-11	MGB2N8	11	1.15	9.71	9.42	96.6	
M1300-12	MGB2N9	12	1.19	9.54	9.39	98.2	
M1300-13	MGB2P0	13	1.18	9.57	8.66	89.2	
M1300-14	MGB2P1	14	1.16	9.83	8.74	87.4	
M1300-15	MGB2P2	15	1.16	9.93	9.72	97.6	
M1300-16	MGB2P3	16	1.14	9.9	9.7	97.7	
M1300-17	MGB2P4	17	1.13	9.7	9.47	97.3	
M1300-18	MGB2P5	18	1.13	9.65	9.48	98.0	
M1300-19	MGB2P6	19	1.16	9.66	9.46	97.6	
M1300-20	MGB2P7	20	1.13	9.72	9.55	98.0	
M1300-21	MGB2P8	21	1.14	9.66	9.46	97.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$