



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Alexander

Parameter: ASTM COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB113033

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP88961
COD calibration std. 100 ppm	WP89077
COD calibration std. 10 ppm	WP89075
COD calibration std. 150 ppm	WP89080
COD calibration std. 50 ppm	WP89076
COD calibration std. 0 ppm	WP89074
COD Digestion Vials Low Level 0-150Mg/L	W2778
COD ICV-LCS std, 50ppm	WP89403
COD CCV std, 50ppm	WP89402

Temp In (C): <u>150</u>	Date In: <u>02/04/2021</u>	Time In: <u>13:00</u>
Temp Out (C): <u>150</u>	Date Out: <u>02/04/2021</u>	Time Out: <u>15:00</u>

Intercept: -0.3627

Slope: 0.9929

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	01/13/2021	17:10
2	CAL2	10	1	Water	9.000	01/13/2021	17:10
3	CAL3	50	1	Water	49.000	01/13/2021	17:11
4	CAL4	100	1	Water	100.000	01/13/2021	17:11
5	CAL5	150	1	Water	148.000	01/13/2021	17:12



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Alexander

Parameter: ASTM COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB113033

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	49.000	49.716	01/13/2021	17:14
2	ICB				1	Water	0.000	0.365	01/13/2021	17:14
3	CCV1	50			1	Water	51.000	51.730	02/04/2021	16:00
4	CCB1				1	Water	0.000	0.365	02/04/2021	16:01
5	LB113033BL				1	Water	0.000	0.365	02/04/2021	16:02
6	LB113033BS	50			1	Water	50.000	50.723	02/04/2021	16:02
7	M1293-03				1	Water	32.000	32.594	02/04/2021	16:03
8	M1293-03DUP				1	Water	31.000	31.587	02/04/2021	16:03
9	M1293-03MS	50			1	Water	78.000	78.923	02/04/2021	16:04
10	M1293-03MSD	50			1	Water	77.000	77.916	02/04/2021	16:05
11	CCV2	50			1	Water	51.000	51.730	02/04/2021	16:05
12	CCB2				1	Water	0.000	0.365	02/04/2021	16:06