



PERCENT SOLID

Supervisor: apatel
Analyst: ketankumar
Date: 2/5/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 16:16
In Date: 02/04/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 105
Time OUT: 10:15
Out Date: 02/05/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB113034

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M1315-01	SP-1	1	1.19	8.96	8.11	89.1	
M1315-02	SP-1A	2	1.18	9.09	8.05	86.9	
M1315-03	TS-1A	3	1.18	9.09	7.67	82.0	
M1315-04	TS-1	4	1.19	9.12	7.68	81.8	
M1324-01	01-A-01-B-01-C	5	1.00	2.00	2.00	100.0	Caulking
M1324-02	02-A-02-B-02-C	6	1.00	2.00	2.00	100.0	Caulking
M1324-03	03-A-03-B-03-C	7	1.00	2.00	2.00	100.0	Caulking

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

LB113034

WorkList Name : 020421_%S WorkList ID : 146153 Department : Wet-Chemistry Date : 02-04-2021 14:21:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/11/2021	Solid	M1315-01	Percent Solids	Cool 4 deg C	EARTH03	G42	SP-1	02/04/2021	Chemtech -SO
02/11/2021	Solid	M1315-02	Percent Solids	Cool 4 deg C	EARTH03	G42	SP-1A	02/04/2021	Chemtech -SO
02/11/2021	Solid	M1315-03	Percent Solids	Cool 4 deg C	EARTH03	G42	TS-1A	02/04/2021	Chemtech -SO
02/11/2021	Solid	M1315-04	Percent Solids	Cool 4 deg C	EARTH03	G42	TS-1	02/04/2021	Chemtech -SO
02/08/2021	Solid	M1324-01	Percent Solids	Cool 4 deg C	BSIG01	G42	01-A-01-B-01-C	02/04/2021	Chemtech -SO
02/08/2021	Solid	M1324-02	Percent Solids	Cool 4 deg C	BSIG01	G42	02-A-02-B-02-C	02/04/2021	Chemtech -SO
02/08/2021	Solid	M1324-03	Percent Solids	Cool 4 deg C	BSIG01	G42	03-A-03-B-03-C	02/04/2021	Chemtech -SO

Date/Time 02/04/21 @ 14:40
 Received by: AKS (Hq-branch) AKS
 Relinquished by: AKS (Hq-branch) AKS

Date/Time 02/04/2021 @ 14:40
 Received by: AKS (Hq-branch)
 Relinquished by: AKS (Hq-branch)