



Analytical Summary Report

Analysis Method: 9040C

Analyst By : ketankumar

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113035

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	02/04/2021	16:05
2	CAL2	1	Water	NA	NA	20.2	7.00	02/04/2021	16:06
3	CAL3	1	Water	NA	NA	20.3	10.04	02/04/2021	16:07
4	ICV	1	Water	NA	NA	20.3	7.00	02/04/2021	16:08
5	CCV1	1	Water	NA	NA	20.2	2.02	02/04/2021	16:09
6	M1313-04	1	Water	NA	NA	10.2	7.94	02/04/2021	16:10
7	M1313-04DUP	1	Water	NA	NA	10.3	7.95	02/04/2021	16:11
8	M1314-04	1	Water	NA	NA	10.7	8.23	02/04/2021	16:12
9	CCV2	1	Water	NA	NA	20.1	12.01	02/04/2021	16:13

LB113035

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH2421 WorkList ID : 146160 Department : Wet-Chemistry Date : 02-04-2021 15:50:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/04/2021	Water	M1313-04	pH	Cool 4 deg C	PSEG03	I31	103	02/04/2021	9040C
02/04/2021	Water	M1314-04	pH	Cool 4 deg C	PSEG03	I31	40966	02/04/2021	9040C

Date/Time 02/04/21 16:00
 Received by: KS (Wetchem)
 Relinquished by: KS (Wetchem)

Date/Time 02/04/21 16:15
 Received by: KS (Wetchem)
 Relinquished by: KS (Wetchem)