


LB113038

7470A

INSTRUMENT 1D: CV1



| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|----|
| 0.0    | 0.000  | 0.022  | 0.022  | 192   | 0.000      | 192   |       |       |       |       | -  |
| 0.2    | 0.200  | 0.258  | 0.058  | 1077  | 0.0 %      | 1077  |       |       |       |       | 29 |
| 2.5    | 2.500  | 2.528  | 0.028  | 9619  | 0.0 %      | 9619  |       |       |       |       | 1  |
| 5.0    | 5.000  | 4.874  | -0.126 | 18448 | 0.0 %      | 18448 |       |       |       |       | -3 |
| 7.5    | 7.500  | 7.376  | -0.124 | 27860 | 0.0 %      | 27860 |       |       |       |       | -2 |
| 10.0   | 10.000 | 10.142 | 0.142  | 38270 | 0.0 %      | 38270 |       |       |       |       | 1  |

## LB113038 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Feb 2021 13:10:22

| Sample ID       | Extended ID | $\mu$ Abs. | Conc.   | Std Conc | Method | Units | Date                 | Type | Type |
|-----------------|-------------|------------|---------|----------|--------|-------|----------------------|------|------|
| 0.0 - 1         | 50          | 192        | -       | 0.0000   | 7470A  | PPB   | 05 Feb 2021 13:52:26 | S    | Std  |
| 0.2 - 1         | 502         | 1077       | -       | 0.2000   | 7470A  | PPB   | 05 Feb 2021 13:54:44 | S    | Std  |
| 2.5 - 1         | 52.5        | 9619       | -       | 2.5000   | 7470A  | PPB   | 05 Feb 2021 13:57:01 | S    | Std  |
| 5.0 - 1         | 55          | 18448      | -       | 5.0000   | 7470A  | PPB   | 05 Feb 2021 13:59:18 | S    | Std  |
| 7.5 - 1         | 57.5        | 27860      | -       | 7.5000   | 7470A  | PPB   | 05 Feb 2021 14:01:34 | S    | Std  |
| 10.0 - 1        | 510         | 38270      | -       | 10.0000  | 7470A  | PPB   | 05 Feb 2021 14:09:13 | S    | Std  |
| ICV92 - 1       | ICV92       | 15349      | 4.0506  | -        | 7470A  | PPB   | 05 Feb 2021 14:15:19 | U    | SMPL |
| ICB92 - 1       | ICB92       | 136        | 0.0075  | -        | 7470A  | PPB   | 05 Feb 2021 14:17:34 | U    | SMPL |
| CCV25 - 1       | CCV25       | 18521      | 4.8937  | -        | 7470A  | PPB   | 05 Feb 2021 14:19:49 | U    | SMPL |
| CCB25 - 1       | CCB25       | 125        | 0.0045  | -        | 7470A  | PPB   | 05 Feb 2021 14:22:04 | U    | SMPL |
| CRA - 1         | CRA         | 981        | 0.2320  | -        | 7470A  | PPB   | 05 Feb 2021 14:24:20 | U    | SMPL |
| HighStd - 1     | HighStd     | 38204      | 10.1248 | -        | 7470A  | PPB   | 05 Feb 2021 14:26:35 | U    | SMPL |
| ChkStd - 1      | ChkStd      | 25295      | 6.6940  | -        | 7470A  | PPB   | 05 Feb 2021 14:28:51 | U    | SMPL |
| PB134484BL - 1  | PBW         | 238        | 0.0346  | -        | 7470A  | PPB   | 05 Feb 2021 14:31:06 | U    | SMPL |
| PB134484BS - 1  | LCSW        | 14283      | 3.7673  | -        | 7470A  | PPB   | 05 Feb 2021 14:33:22 | U    | SMPL |
| M1313-03 - 1    | 30743       | 13905      | 3.6669  | -        | 7470A  | PPB   | 05 Feb 2021 14:35:39 | U    | SMPL |
| M1313-03DUP - 1 | 30743DUP    | 13478      | 3.5534  | -        | 7470A  | PPB   | 05 Feb 2021 14:37:55 | U    | SMPL |
| M1313-03MS - 1  | 30743MS     | 27636      | 7.3162  | -        | 7470A  | PPB   | 05 Feb 2021 14:40:12 | U    | SMPL |
| M1313-03MSD - 1 | 30743MSD    | 27056      | 7.1620  | -        | 7470A  | PPB   | 05 Feb 2021 14:42:30 | U    | SMPL |
| CCV26 - 1       | CCV26       | 18418      | 4.8663  | -        | 7470A  | PPB   | 05 Feb 2021 14:44:45 | U    | SMPL |
| CCB26 - 1       | CCB26       | 96         | -0.0032 | -        | 7470A  | PPB   | 05 Feb 2021 14:47:01 | U    | SMPL |
| PB134484TB - 1  | PB134470TB  | 294        | 0.0494  | -        | 7470A  | PPB   | 05 Feb 2021 14:49:17 | U    | SMPL |
| M1313-03LX5 - 1 |             | 2951       | 0.7556  | -        | 7470A  | PPB   | 05 Feb 2021 14:51:32 | U    | SMPL |
| M1313-03A - 1   |             | 27053      | 7.1612  | -        | 7470A  | PPB   | 05 Feb 2021 14:53:48 | U    | SMPL |
| CCV27 - 1       | CCV27       | 18594      | 4.9131  | -        | 7470A  | PPB   | 05 Feb 2021 14:56:03 | U    | SMPL |
| CCB27 - 1       | CCB27       | 121        | 0.0035  | -        | 7470A  | PPB   | 05 Feb 2021 14:58:19 | U    | SMPL |