



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB113042

Analyst By : ketankumar

Supervisor Review By : apatel

Slope : 98.7

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	02/05/2021	13:20
2	CAL2	1	Water	NA	NA	20.3	7.00	02/05/2021	13:21
3	CAL3	1	Water	NA	NA	20.1	10.02	02/05/2021	13:22
4	ICV	1	Water	NA	NA	2.02	7.00	02/05/2021	13:23
5	CCV1	1	Water	NA	NA	20.1	2.02	02/05/2021	13:24
6	M1316-09	1	Water	NA	NA	20.3	7.83	02/05/2021	13:25
7	M1316-09DUP	1	Water	NA	NA	20.4	7.84	02/05/2021	13:26
8	M1316-12	1	Water	NA	NA	20.1	7.90	02/05/2021	13:27
9	M1316-13	1	Water	NA	NA	20.5	7.88	02/05/2021	13:28
10	M1316-14	1	Water	NA	NA	20.4	7.99	02/05/2021	13:29
11	M1318-02	1	Water	NA	NA	20.3	5.79	02/05/2021	13:30
12	M1318-05	1	Water	NA	NA	20.2	5.68	02/05/2021	13:31
13	M1318-06	1	Water	NA	NA	20.3	5.81	02/05/2021	13:32
14	M1318-07	1	Water	NA	NA	20.2	5.69	02/05/2021	13:33
15	M1318-09	1	Water	NA	NA	20.6	5.92	02/05/2021	13:34
16	CCV2	1	Water	NA	NA	20.2	12.01	02/05/2021	13:35
17	M1318-10	1	Water	NA	NA	20.4	5.59	02/05/2021	13:36
18	M1318-11	1	Water	NA	NA	20.3	5.23	02/05/2021	13:37
19	M1318-12	1	Water	NA	NA	20.5	5.64	02/05/2021	13:38
20	CCV3	1	Water	NA	NA	20.1	2.01	02/05/2021	13:39

WORKLIST(Hardcopy Internal Chain)

LB113042

WorkList Name : ph2521-1 WorkList ID : 146190 Department : Wet-Chemistry Date : 02-05-2021 13:05:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/04/2021	Water	M1316-09	pH	Cool 4 deg C	PSEG03	I32	OILY-WATER-TOTES-COMP	02/04/2021	9040C
02/04/2021	Water	M1316-12	pH	Cool 4 deg C	PSEG03	I32	31417	02/04/2021	9040C
02/04/2021	Water	M1316-13	pH	Cool 4 deg C	PSEG03	I32	31412	02/04/2021	9040C
02/04/2021	Water	M1316-14	pH	Cool 4 deg C	PSEG03	I32	31416	02/04/2021	9040C
02/04/2021	Water	M1318-02	pH	Cool 4 deg C	PSEG03	I31	31408-31410	02/04/2021	9040C
02/04/2021	Water	M1318-05	pH	Cool 4 deg C	PSEG03	I31	31408	02/04/2021	9040C
02/04/2021	Water	M1318-06	pH	Cool 4 deg C	PSEG03	I31	31409	02/04/2021	9040C
02/04/2021	Water	M1318-07	pH	Cool 4 deg C	PSEG03	I31	31410	02/04/2021	9040C
02/04/2021	Water	M1318-09	pH	Cool 4 deg C	PSEG03	I31	31394	02/04/2021	9040C
02/04/2021	Water	M1318-10	pH	Cool 4 deg C	PSEG03	I31	31395	02/04/2021	9040C
02/04/2021	Water	M1318-11	pH	Cool 4 deg C	PSEG03	I31	31403	02/04/2021	9040C
02/04/2021	Water	M1318-12	pH	Cool 4 deg C	PSEG03	I31	31387	02/04/2021	9040C

Date/Time 02/05/21 13:15
 Received by: KS (Weber)
 Relinquished by: KS (Weber)

Date/Time 02/05/21 13:45
 Received by: KS (Weber)
 Relinquished by: KS (Weber)