



Analytical Summary Report

Analysis Method: 9045D

Analyst By : ketankumar

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113044

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.02	02/05/2021	13:55
2	CAL2	1	Water	NA	NA	20.3	7.00	02/05/2021	13:56
3	CAL3	1	Water	NA	NA	20.1	10.02	02/05/2021	13:57
4	ICV	1	Water	NA	NA	20.1	7.00	02/05/2021	13:58
5	CCV1	1	Water	NA	NA	20.3	2.02	02/05/2021	13:59
6	M1316-01	1	Solid	20	20	20.2	5.91	02/05/2021	14:00
7	M1316-01DUP	1	Solid	20	20	20.3	5.92	02/05/2021	14:01
8	M1316-04	1	Solid	20	20	20.1	5.97	02/05/2021	14:02
9	M1316-05	1	Solid	20	20	20.3	6.05	02/05/2021	14:03
10	M1316-06	1	Solid	20	20	20.2	6.40	02/05/2021	14:04
11	CCV2	1	Water	NA	NA	20.2	12.01	02/05/2021	14:05

WORKLIST(Hardcopy Internal Chain)

LB113049

WorkList Name : ph2521

WorkList ID : 146189

Department : Wet-Chemistry

Date : 02-05-2021 13:04:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/04/2021	Solid	M1316-01	pH	Cool 4 deg C	PSEG03	I32	OILY-DEBRIS-DRUMS-COMF	02/04/2021	9045D
02/04/2021	Solid	M1316-04	pH	Cool 4 deg C	PSEG03	I32	31414	02/04/2021	9045D
02/04/2021	Solid	M1316-05	pH	Cool 4 deg C	PSEG03	I32	31415	02/04/2021	9045D
02/04/2021	Solid	M1316-06	pH	Cool 4 deg C	PSEG03	I32	31413	02/04/2021	9045D

Date/Time 02/05/21 13:50
 Received by: KS (Welcher)
 Relinquished by: KS (Welcher)

Date/Time 02/05/21 14:10
 Received by: KS (Welcher)
 Relinquished by: KS (Welcher)