


LB113045

7471B

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 1.7976e-004

C= -7.5215e-002

Rho= 0.9998663

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%b
0.0	0.000	-0.024	-0.024	286	0.000	286					-
0.2	0.200	0.187	-0.013	1461	0.0 %	1461					-7
2.5	2.500	2.468	-0.032	14145	0.0 %	14145					-1
5.0	5.000	5.096	0.096	28765	0.0 %	28765					2
7.5	7.500	7.559	0.059	42469	0.0 %	42469					1
10.0	10.000	9.914	-0.086	55568	0.0 %	55568					-1

LB113045 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Feb 2021 15:44:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	286	-	0.0000	7471B	PPB	05 Feb 2021 15:45:09	S	Std
0.2 - 1	50.2	1461	-	0.2000	7471B	PPB	05 Feb 2021 15:47:26	S	Std
2.5 - 1	52.5	14145	-	2.5000	7471B	PPB	05 Feb 2021 15:49:44	S	Std
5.0 - 1	55	28765	-	5.0000	7471B	PPB	05 Feb 2021 15:52:01	S	Std
7.5 - 1	57.5	42469	-	7.5000	7471B	PPB	05 Feb 2021 15:54:18	S	Std
10.0 - 1	510	55568	-	10.0000	7471B	PPB	05 Feb 2021 15:56:34	S	Std
ICV93 - 1	ICV93	22561	3.9804	-7471B	PPB	PPB	05 Feb 2021 16:00:46	U	SMPL
ICB93 - 1	ICB93	246	-0.0310	-7471B	PPB	PPB	05 Feb 2021 16:06:33	U	SMPL
CCV28 - 1	CCV28	28844	5.1099	-7471B	PPB	PPB	05 Feb 2021 16:08:48	U	SMPL
CCB28 - 1	CCB28	149	-0.0484	-7471B	PPB	PPB	05 Feb 2021 16:11:03	U	SMPL
CRA - 1	CRA	1432	0.1822	-7471B	PPB	PPB	05 Feb 2021 16:13:19	U	SMPL
HighStd - 1	HighStd	55816	9.9585	-7471B	PPB	PPB	05 Feb 2021 16:15:34	U	SMPL
ChkStd - 1	ChkStd	42659	7.5934	-7471B	PPB	PPB	05 Feb 2021 16:23:04	U	SMPL
PB134486BL - 1	PBS	126	-0.0526	-7471B	PPB	PPB	05 Feb 2021 16:25:19	U	SMPL
PB134486BS - 1	LCSS	22318	3.9368	-7471B	PPB	PPB	05 Feb 2021 16:27:36	U	SMPL
M1313-01 - 1	41710	424	0.0010	-7471B	PPB	PPB	05 Feb 2021 16:29:53	U	SMPL
M1314-03 - 1	30752	1243	0.1482	-7471B	PPB	PPB	05 Feb 2021 16:32:09	U	SMPL
M1315-01 - 1	SP-1	968	0.0988	-7471B	PPB	PPB	05 Feb 2021 16:34:26	U	SMPL
M1315-04 - 1	TS-1	3910	0.6277	-7471B	PPB	PPB	05 Feb 2021 16:36:43	U	SMPL
M1315-04DUP - 1	TS-1DUP	3614	0.5745	-7471B	PPB	PPB	05 Feb 2021 16:41:35	U	SMPL
CCV29 - 1	CCV29	26837	4.7491	-7471B	PPB	PPB	05 Feb 2021 16:43:50	U	SMPL
CCB29 - 1	CCB29	129	-0.0520	-7471B	PPB	PPB	05 Feb 2021 16:46:05	U	SMPL
M1315-04MS - 1	TS-1MS	25765	4.5564	-7471B	PPB	PPB	05 Feb 2021 16:48:21	U	SMPL
M1315-04MSD - 1	TS-1MSD	24421	4.3148	-7471B	PPB	PPB	05 Feb 2021 16:50:37	U	SMPL
M1317-01 - 1	R-A	3277	0.5139	-7471B	PPB	PPB	05 Feb 2021 16:52:52	U	SMPL
M1317-03 - 1	R-B	29981	5.3143	-7471B	PPB	PPB	05 Feb 2021 16:55:08	U	SMPL
M1317-05 - 1	R-C	6163	1.0327	-7471B	PPB	PPB	05 Feb 2021 16:57:24	U	SMPL
M1315-04LX5 - 1		855	0.0785	-7471B	PPB	PPB	05 Feb 2021 16:59:41	U	SMPL
M1315-04A - 1		25984	4.5958	-7471B	PPB	PPB	05 Feb 2021 17:01:58	U	SMPL
CCV30 - 1	CCV30	27008	4.7799	-7471B	PPB	PPB	05 Feb 2021 17:04:14	U	SMPL
CCB30 - 1	CCB30	143	-0.0495	-7471B	PPB	PPB	05 Feb 2021 17:06:31	U	SMPL