



Analytical Summary Report

Analysis Method: 9045D

Analyst By : ketankumar

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB113055

Slope : 98.7

BalanceID:

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.00	02/06/2021	09:45
2	CAL2	1	Water	NA	NA	20.2	7.01	02/06/2021	09:46
3	CAL3	1	Water	NA	NA	20.1	10.02	02/06/2021	09:47
4	ICV	1	Water	NA	NA	20.2	7.00	02/06/2021	09:48
5	CCV1	1	Water	NA	NA	20.1	2.02	02/06/2021	09:49
6	M1317-11	1	Solid	20	20	20.4	6.95	02/06/2021	09:50
7	M1317-11DUP	1	Solid	20	20	20.5	6.97	02/06/2021	09:51
8	M1317-13	1	Solid	20	20	20.6	6.72	02/06/2021	09:52
9	M1317-15	1	Solid	20	20	20.5	7.78	02/06/2021	09:53
10	CCV2	1	Water	NA	NA	20.2	12.01	02/06/2021	09:54

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORO2621 WorkList ID : 146203 Department : Wet-Chemistry Date : 02-06-2021 09:32:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/04/2021	Solid	M1317-11	Corrosivity	Cool 4 deg C	PSEG03	I32	R-A	02/04/2021	9045D
02/04/2021	Solid	M1317-13	Corrosivity	Cool 4 deg C	PSEG03	I32	R-B	02/04/2021	9045D
02/04/2021	Solid	M1317-15	Corrosivity	Cool 4 deg C	PSEG03	I32	R-C	02/04/2021	9045D

Date/Time 02/06/21 09:35
 Received by: KS (Wetchem)
 Relinquished by: KS (Wetchem)

Date/Time 02/06/21 19:00
 Received by: KS (Wetchem)
 Relinquished by: KS (Wetchem)