



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB113063

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: FLASHPONT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.9	02/06/2021	09:25
2	M1313-01		1	9	100.00	>212.0	>212.0	02/06/2021	09:50
3	M1313-01DUP		1	9	100.00	>212.0	>212.0	02/06/2021	10:15
4	M1314-01		1	11	100.00	>212.0	>212.0	02/06/2021	10:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 113063

WorkList Name : FP2621

WorkList ID : 146201

Department : Wet-Chemistry

Date : 02-06-2021 08:47:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/11/2021	Solid	M1313-01	Flash Point	Cool 4 deg C	PSEG03	I31	41710	02/04/2021	1010A
02/11/2021	Solid	M1314-01	Flash Point	Cool 4 deg C	PSEG03	I31	40966	02/04/2021	1010A

Date/Time 02/06/21 08:55
 Received by: KS (Wetchem)
 Relinquished by: KS (Wetchem)

Date/Time 02/06/21 10:45
 Received by: KS (Wetchem)
 Relinquished by: KS (Wetchem)