



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB113064

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: FLASHPONT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.9	02/06/2021	09:25
2	M1313-04		1	10	100.00	>212.0	>212.0	02/06/2021	11:15
3	M1313-04DUP		1	10	100.00	>212.0	>212.0	02/06/2021	11:40
4	M1314-04		1	9	100.00	>212.0	>212.0	02/06/2021	12:05
5	M1316-09		1	9	100.00	>212.0	>212.0	02/06/2021	12:35
6	M1318-02		1	10	100.00	>212.0	>212.0	02/06/2021	13:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 113064

WorkList Name : FP2621-1 WorkList ID : 146202 Department : Wet-Chemistry Date : 02-06-2021 08:48:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/11/2021	Water	M1313-04	C Flash Point	Cool 4 deg C	PSEG03	I31	103	02/04/2021	1010A
02/11/2021	Water	M1314-04	F Flash Point	Cool 4 deg C	PSEG03	I31	40966	02/04/2021	1010A
02/09/2021	Water	M1316-09	C Flash Point	Cool 4 deg C	PSEG03	I32	OILY-WATER-TOTES-COMP	02/04/2021	1010A
02/18/2021	Water	M1318-02	C Flash Point	Cool 4 deg C	PSEG03	I31	31408-31410	02/04/2021	1010A

Date/Time 02/06/21 10:50
 Received by: KS (Welcher)
 Relinquished by: KS (Welcher)

Date/Time 02/06/21 13:15
 Received by: KS (Welcher)
 Relinquished by: KS (Welcher)