



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : ketankumar

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113065

Slope : 98.7

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	02/08/2021	15:02
2	CAL2	1	Water	NA	NA	20.2	7.00	02/08/2021	15:03
3	CAL3	1	Water	NA	NA	20.1	10.02	02/08/2021	15:04
4	ICV	1	Water	NA	NA	20.1	7.00	02/08/2021	15:05
5	CCV1	1	Water	NA	NA	20.2	2.02	02/08/2021	15:06
6	M1336-01	1	Water	NA	NA	20.1	4.91	02/08/2021	15:07
7	M1336-01DUP	1	Water	NA	NA	20.2	4.92	02/08/2021	15:08
8	CCV2	1	Water	NA	NA	20.1	12.01	02/08/2021	15:09

WORKLIST(Hardcopy Internal Chain)

LB 113065

WorkList Name : PH2821 WorkList ID : 146227 Department : Wet-Chemistry Date : 02-08-2021 14:59:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/04/2021	Water	M1336-01	pH	Cool 4 deg C	STER04	G61	POOL-WATER	02/04/2021	SM 4500-H B

Date/Time 02/08/21 15:00
 Received by: KS (Weber)
 Relinquished by: KS (Weber)

Date/Time 02/08/21 15:10
 Received by: KS (Weber)
 Relinquished by: KS (Weber)