


LA113066

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0 D
0.0	0.000	0.009	0.009	65	0.000	65					-
0.2	0.200	0.221	0.021	852	0.0 %	852					11
2.5	2.500	2.319	-0.181	8622	0.0 %	8622					-7
5.0	5.000	5.161	0.161	19144	0.0 %	19144					3
7.5	7.500	7.643	0.143	28335	0.0 %	28335					2
10.0	10.000	9.847	-0.153	36497	0.0 %	36497					-2

LB113066 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 08 Feb 2021 15:26:52				
Sample ID	Extended ID	µ Abs	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50	65	-	0.0000	7470A	PPB	08 Feb 2021 15:27:14	S Std
0.2 - 1	50.2	852	-	0.2000	7470A	PPB	08 Feb 2021 15:29:32	S Std
2.5 - 1	52.5	8622	-	2.5000	7470A	PPB	08 Feb 2021 15:31:49	S Std
5.0 - 1	55	19144	-	5.0000	7470A	PPB	08 Feb 2021 15:34:06	S Std
7.5 - 1	57.5	28335	-	7.5000	7470A	PPB	08 Feb 2021 15:36:23	S Std
10.0 - 1	518	36497	-	10.0000	7470A	PPB	08 Feb 2021 15:38:38	S Std
ICV95 - 1	ICV95	13651	3.6775	-	7470A	PPB	08 Feb 2021 15:41:38	U SMPL
ICB95 - 1	ICB95	-48	-0.0218	-	7470A	PPB	08 Feb 2021 15:43:53	U SMPL
CCV38 - 1	CCV38	19232	5.1846	-	7470A	PPB	08 Feb 2021 15:46:08	U SMPL
CCB38 - 1	CCB38	-45	-0.0210	-	7470A	PPB	08 Feb 2021 15:48:24	U SMPL
CRA - 1	CRA	807	0.2091	-	7470A	PPB	08 Feb 2021 15:50:39	U SMPL
HighStd - 1	HighStd	37235	10.0462	-	7470A	PPB	08 Feb 2021 15:55:54	U SMPL
ChkStd - 1	ChkStd	26407	7.1222	-	7470A	PPB	08 Feb 2021 15:58:09	U SMPL
PB134508BL - 1	PBW	-82	-0.0310	-	7470A	PPB	08 Feb 2021 16:00:24	U SMPL
PB134508BS - 1	LCSW	13728	3.6983	-	7470A	PPB	08 Feb 2021 16:02:41	U SMPL
M1316-02 - 1	OILY-DEBRIS-DRUMS-COMP	-58	-0.0245	-	7470A	PPB	08 Feb 2021 16:04:57	U SMPL
M1316-02DUP - 1	OILY-DEBRIS-DRUMS-COMPDUP	-3	-0.0096	-	7470A	PPB	08 Feb 2021 16:07:14	U SMPL
M1316-02MS - 1	OILY-DEBRIS-DRUMS-COMPMS	15185	4.0918	-	7470A	PPB	08 Feb 2021 16:09:31	U SMPL
M1316-02MSD - 1	OILY-DEBRIS-DRUMS-COMPMSD	13441	3.6208	-	7470A	PPB	08 Feb 2021 16:11:48	U SMPL
M1317-11 - 1	R-A	-13	-0.0123	-	7470A	PPB	08 Feb 2021 16:14:03	U SMPL
CCV39 - 1	CCV39	19298	5.2024	-	7470A	PPB	08 Feb 2021 16:16:19	U SMPL
CCB39 - 1	CCB39	-49	-0.0221	-	7470A	PPB	08 Feb 2021 16:18:35	U SMPL
M1317-13 - 1	R-B	12	-0.0056	-	7470A	PPB	08 Feb 2021 16:20:50	U SMPL
M1317-15 - 1	R-C	9	-0.0064	-	7470A	PPB	08 Feb 2021 16:23:06	U SMPL
PB134509BL - 1	PBW	4	-0.0077	-	7470A	PPB	08 Feb 2021 16:25:21	U SMPL
PB134509BS - 1	LCSW	14830	3.9959	-	7470A	PPB	08 Feb 2021 16:27:37	U SMPL
M1316-10 - 1	OILY-WATER-TOTES-COMP	-20	-0.0142	-	7470A	PPB	08 Feb 2021 16:29:53	U SMPL
M1316-10DUP - 1	OILY-WATER-TOTES-COMPDUP	12	-0.0056	-	7470A	PPB	08 Feb 2021 16:32:10	U SMPL
M1316-10MS - 1	OILY-WATER-TOTES-COMPMS	16315	4.3969	-	7470A	PPB	08 Feb 2021 16:34:26	U SMPL
M1316-10MSD - 1	OILY-WATER-TOTES-COMPMSD	16259	4.3818	-	7470A	PPB	08 Feb 2021 16:36:43	U SMPL
M1318-03 - 1	31408-31410	464	0.1165	-	7470A	PPB	08 Feb 2021 16:39:00	U SMPL
M1333-02 - 1	CHRT27339-CHRT20101	-20	-0.0142	-	7470A	PPB	08 Feb 2021 16:41:16	U SMPL
CCV40 - 1	CCV40	19517	5.2616	-	7470A	PPB	08 Feb 2021 16:43:32	U SMPL
CCB40 - 1	CCB40	-34	-0.0180	-	7470A	PPB	08 Feb 2021 16:45:47	U SMPL
PB134485BL - 1	PBW	5	-0.0075	-	7470A	PPB	08 Feb 2021 16:50:47	U SMPL
PB134485BS - 1	LCSW	13135	3.5382	-	7470A	PPB	08 Feb 2021 16:53:02	U SMPL
M1291-20 - 1	2021-01-27-EB	235	0.0546	-	7470A	PPB	08 Feb 2021 16:55:18	U SMPL
M1292-18 - 1	2021-01-28-EB	263	0.0622	-	7470A	PPB	08 Feb 2021 16:57:34	U SMPL
M1314-04 - 1	40966	546	0.1386	-	7470A	PPB	08 Feb 2021 16:59:50	U SMPL
M1333-01 - 1	CHRT27339-CHRT20101	168	0.0365	-	7470A	PPB	08 Feb 2021 17:02:06	U SMPL
M1333-01DUP - 1	CHRT27339-CHRT20101DUP	156	0.0333	-	7470A	PPB	08 Feb 2021 17:04:23	U SMPL
M1333-01MS - 1	CHRT27339-CHRT20101MS	15244	4.1077	-	7470A	PPB	08 Feb 2021 17:06:40	U SMPL
M1333-01MSD - 1	CHRT27339-CHRT20101MSD	15284	4.1185	-	7470A	PPB	08 Feb 2021 17:08:57	U SMPL
CCV41 - 1	CCV41	18420	4.9653	-	7470A	PPB	08 Feb 2021 17:13:55	U SMPL
CCB41 - 1	CCB41	-84	-0.0315	-	7470A	PPB	08 Feb 2021 17:16:10	U SMPL
PB134508TB - 1	PB134477TB	-13	-0.0123	-	7470A	PPB	08 Feb 2021 17:18:26	U SMPL
PB134509TB - 1	PB134495TB	-3	-0.0096	-	7470A	PPB	08 Feb 2021 17:20:41	U SMPL
M1316-02LX5 - 1		4	-0.0077	-	7470A	PPB	08 Feb 2021 17:22:57	U SMPL
M1316-02A - 1		14821	3.9935	-	7470A	PPB	08 Feb 2021 17:25:13	U SMPL
M1316-10LX5 - 1		-40	-0.0196	-	7470A	PPB	08 Feb 2021 17:27:29	U SMPL
M1316-10A - 1		17106	4.6105	-	7470A	PPB	08 Feb 2021 17:29:45	U SMPL
M1333-01LX5 - 1		-17	-0.0134	-	7470A	PPB	08 Feb 2021 17:32:02	U SMPL
M1333-01A - 1		15561	4.1933	-	7470A	PPB	08 Feb 2021 17:34:18	U SMPL
CCV42 - 1	CCV42	19482	5.2521	-	7470A	PPB	08 Feb 2021 17:36:35	U SMPL
CCB42 - 1	CCB42	-65	-0.0264	-	7470A	PPB	08 Feb 2021 17:38:52	U SMPL