

LB 113078


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

59517



A= 0.0000e+000

B= 1.6834e-004

C= 2.8009e-003

Rho= 0.9999915

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%ob
0.0	0.000	0.001	0.001	-11	0.000	-11					1
0.2	0.200	0.218	0.018	1278	0.0 %	1278					9
2.5	2.500	2.492	-0.008	14787	0.0 %	14787					0
5.0	5.000	4.983	-0.017	29585	0.0 %	29585					0
7.5	7.500	7.484	-0.016	44444	0.0 %	44444					0
10.0	10.000	10.022	0.022	59517	0.0 %	59517					0

LB113073 INSTRUMENT ID : CV1

Method: 7471B	Operator: Admin	Date of Analysis: 09 Feb 2021 12:37:44					Date	Type	Type
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units			
0.0 - 1	SC	-11	-	0.00007471B	PPB		09 Feb 2021 12:39:22	S	Std
0.2 - 1	SC.2	1278	-	0.20007471B	PPB		09 Feb 2021 12:41:39	S	Std
2.5 - 1	SC.5	14787	-	2.50007471B	PPB		09 Feb 2021 12:43:57	S	Std
5.0 - 1	SS	29585	-	5.00007471B	PPB		09 Feb 2021 12:46:13	S	Std
7.5 - 1	SC.5	44444	-	7.50007471B	PPB		09 Feb 2021 12:48:30	S	Std
10.0 - 1	SIU	59517	-	10.00007471B	PPB		09 Feb 2021 12:56:04	S	Std
ICV96 - 1	ICV96	22467	3.7848	-7471B	PPB		09 Feb 2021 12:58:42	U	SMPL
ICB96 - 1	ICB96	-65	-0.0081	-7471B	PPB		09 Feb 2021 13:00:56	U	SMPL
CCV43 - 1	CCV43	29377	4.9480	-7471B	PPB		09 Feb 2021 13:03:11	U	SMPL
CCB43 - 1	CCB43	-74	-0.0097	-7471B	PPB		09 Feb 2021 13:05:27	U	SMPL
CRA - 1	CRA	1270	0.2166	-7471B	PPB		09 Feb 2021 13:07:42	U	SMPL
HighStd - 1	HighStd	59880	10.0828	-7471B	PPB		09 Feb 2021 13:09:57	U	SMPL
ChkStd - 1	ChkStd	41524	6.9928	-7471B	PPB		09 Feb 2021 13:12:13	U	SMPL
PB134539BL - 1	PBS	-112	-0.0161	-7471B	PPB		09 Feb 2021 13:14:30	U	SMPL
PB134539BS - 1	LCSS	20122	3.3901	-7471B	PPB		09 Feb 2021 13:16:48	U	SMPL
M1337-02 - 1	CLASSIC_BH_01_0.5-2	3068	0.5193	-7471B	PPB		09 Feb 2021 13:19:04	U	SMPL
M1337-02DUP - 1	CLASSIC_BH_01_0.5-2DUP	2974	0.5034	-7471B	PPB		09 Feb 2021 13:24:02	U	SMPL
M1337-02MS - 1	CLASSIC_BH_01_0.5-2MS	24652	4.1526	-7471B	PPB		09 Feb 2021 13:26:18	U	SMPL
M1337-02MSD - 1	CLASSIC_BH_01_0.5-2MSD	25027	4.2158	-7471B	PPB		09 Feb 2021 13:28:35	U	SMPL
M1337-04 - 1	CLASSIC_BH_01_2-4	1244	0.2122	-7471B	PPB		09 Feb 2021 13:30:51	U	SMPL
CCV44 - 1	CCV44	29589	4.9837	-7471B	PPB		09 Feb 2021 13:33:06	U	SMPL
CCB44 - 1	CCB44	-57	-0.0068	-7471B	PPB		09 Feb 2021 13:35:21	U	SMPL
M1339-02 - 1	TP-0-5	4558	0.7701	-7471B	PPB		09 Feb 2021 13:37:37	U	SMPL
M1339-07 - 1	TP-5-10	745	0.1282	-7471B	PPB		09 Feb 2021 13:39:52	U	SMPL
M1340-01 - 1	CHRT-26586	514	0.0893	-7471B	PPB		09 Feb 2021 13:42:07	U	SMPL
M1341-01 - 1	20057	3852	0.6512	-7471B	PPB		09 Feb 2021 13:44:23	U	SMPL
M1341-02 - 1	26572	3948	0.6674	-7471B	PPB		09 Feb 2021 13:46:39	U	SMPL
M1337-02LX5 - 1		499	0.0868	-7471B	PPB		09 Feb 2021 13:48:55	U	SMPL
M1337-02A - 1		25945	4.3703	-7471B	PPB		09 Feb 2021 13:51:12	U	SMPL
CCV45 - 1	CCV45	29502	4.9691	-7471B	PPB		09 Feb 2021 13:53:29	U	SMPL
CCB45 - 1	CCB45	-76	-0.0100	-7471B	PPB		09 Feb 2021 13:55:47	U	SMPL