


LB113074

ISM02.4

INSTRUMENTS ID: EV1

 μ Abs.:

50534



Linear

A= 0.0000e+000

B= 1.9733e-004 *slope*C= -1.5891e-002 *-1.5891e-002 = intercept*

Rho= 0.9999558

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.006	-0.006	52	0.000	52					1
0.05	0.050					334					6
0.20	0.200	0.211	0.011	1151	0.0 %	1151					1
2.50	2.500	2.469	-0.031	12594	0.0 %	12594					0
5.00	5.000	5.004	0.004	25441	0.0 %	25441					0
7.50	7.500	7.565	0.065	38415	0.0 %	38415					1
10.0	10.000	9.956	-0.044	50534	0.0 %	50534					0

LB113074
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	52	-		0 ISM02.4	PPB	2/9/2021 14:57	Std
	0.2 S0.2	1151	-		0.2 ISM02.4	PPB	2/9/2021 14:59	Std
	2.5 S2.5	12594	-		2.5 ISM02.4	PPB	2/9/2021 15:01	Std
	5 S5.0	25441	-		5 ISM02.4	PPB	2/9/2021 15:04	Std
	7.5 S7.5	38415	-		7.5 ISM02.4	PPB	2/9/2021 15:06	Std
	10 S10.0	50534	-		10 ISM02.4	PPB	2/9/2021 15:08	Std
ICV	ICV	19603	3.8524	-	ISM02.4	PPB	2/9/2021 15:11	SMPL
ICB	ICB	-31	-0.022	-	ISM02.4	PPB	2/9/2021 15:13	SMPL
CCV046	CCV046	25364	4.9892	-	ISM02.4	PPB	2/9/2021 15:15	SMPL
CCB046	CCB046	-20	-0.0198	-	ISM02.4	PPB	2/9/2021 15:18	SMPL
PB134540BL	PBS540	19	-0.0121	-	ISM02.4	PPB	2/9/2021 15:20	SMPL
M1307-01	MCOA16	4512	0.8745	-	ISM02.4	PPB	2/9/2021 15:22	SMPL
M1307-02	MCOA17	971	0.1757	-	ISM02.4	PPB	2/9/2021 15:24	SMPL
M1307-03	MCOA18	870	0.1558	-	ISM02.4	PPB	2/9/2021 15:27	SMPL
M1307-04	MCOA19	378	0.0587	-	ISM02.4	PPB	2/9/2021 15:29	SMPL
M1307-05	MCOA20	1419	0.2641	-	ISM02.4	PPB	2/9/2021 15:31	SMPL
M1307-06	MCOA20D	1316	0.2438	-	ISM02.4	PPB	2/9/2021 15:33	SMPL
M1307-07	MCOA20S	14746	2.894	-	ISM02.4	PPB	2/9/2021 15:36	SMPL
M1307-08	MCOA21	843	0.1505	-	ISM02.4	PPB	2/9/2021 15:38	SMPL
M1307-09	MCOA22	762	0.1345	-	ISM02.4	PPB	2/9/2021 15:40	SMPL
M1307-10	MCOA23	1046	0.1905	-	ISM02.4	PPB	2/9/2021 15:43	SMPL
M1307-11	MCOA27	2046	0.3878	-	ISM02.4	PPB	2/9/2021 15:45	SMPL
M1307-12	MCOA29	2193	0.4169	-	ISM02.4	PPB	2/9/2021 15:47	SMPL
M1307-13	MCOA30	2373	0.4524	-	ISM02.4	PPB	2/9/2021 15:49	SMPL
M1307-14	MCOA35	1781	0.3356	-	ISM02.4	PPB	2/9/2021 15:52	SMPL
M1307-15	MCOA36	896	0.1609	-	ISM02.4	PPB	2/9/2021 15:54	SMPL
M1307-16	MCOA37	800	0.142	-	ISM02.4	PPB	2/9/2021 15:56	SMPL
M1307-17	MCOA38	1176	0.2162	-	ISM02.4	PPB	2/9/2021 15:58	SMPL
M1307-18	MCOA40	525	0.0877	-	ISM02.4	PPB	2/9/2021 16:01	SMPL
M1307-19	MCOA41	723	0.1268	-	ISM02.4	PPB	2/9/2021 16:03	SMPL
CCV047	CCV047	25911	5.0972	-	ISM02.4	PPB	2/9/2021 16:05	SMPL
CCB047	CCB047	-20	-0.0198	-	ISM02.4	PPB	2/9/2021 16:08	SMPL
M1307-20	MCOA42	895	0.1607	-	ISM02.4	PPB	2/9/2021 16:10	SMPL
M1307-21	MCOAA2	264	0.0362	-	ISM02.4	PPB	2/9/2021 16:12	SMPL
M1307-22	MCOAA3	282	0.0398	-	ISM02.4	PPB	2/9/2021 16:14	SMPL
PB134541BL	PBS541	47	-0.0066	-	ISM02.4	PPB	2/9/2021 16:17	SMPL
M1308-01	MCOAA4	606	0.1037	-	ISM02.4	PPB	2/9/2021 16:19	SMPL
M1308-02	MCOAA5	1394	0.2592	-	ISM02.4	PPB	2/9/2021 16:21	SMPL
M1308-03	MCOAA6	907	0.1631	-	ISM02.4	PPB	2/9/2021 16:23	SMPL
M1308-04	MCOAA7	2770	0.5307	-	ISM02.4	PPB	2/9/2021 16:26	SMPL
M1308-05	MCOAA7D	2456	0.4688	-	ISM02.4	PPB	2/9/2021 16:28	SMPL
M1308-06	MCOAA7S	15687	3.0796	-	ISM02.4	PPB	2/9/2021 16:30	SMPL
M1308-07	MCOAA8	560	0.0946	-	ISM02.4	PPB	2/9/2021 16:32	SMPL
M1308-08	MCOAA9	454	0.0737	-	ISM02.4	PPB	2/9/2021 16:35	SMPL
M1308-09	MCOAB0	512	0.0851	-	ISM02.4	PPB	2/9/2021 16:37	SMPL
M1308-10	MCOAB1	708	0.1238	-	ISM02.4	PPB	2/9/2021 16:39	SMPL

LB113074
INSTRUMENT ID : CV1

M1308-11	MC0AB2	642	0.1108 -	ISM02.4	PPB	2/9/2021 16:42	SMPL
M1308-12	MC0AB3	548	0.0922 -	ISM02.4	PPB	2/9/2021 16:44	SMPL
M1308-13	MC0AB7	213	0.0261 -	ISM02.4	PPB	2/9/2021 16:46	SMPL
M1308-14	MC0A10	456	0.0741 -	ISM02.4	PPB	2/9/2021 16:48	SMPL
M1308-15	MC0A11	1086	0.1984 -	ISM02.4	PPB	2/9/2021 16:51	SMPL
M1308-16	MC0A12	1064	0.1941 -	ISM02.4	PPB	2/9/2021 16:53	SMPL
CCV048	CCV048	25920	5.0989 -	ISM02.4	PPB	2/9/2021 16:55	SMPL
CCB048	CCB048	-39	-0.0236 -	ISM02.4	PPB	2/9/2021 16:57	SMPL
M1308-17	MC0A13	1156	0.2122 -	ISM02.4	PPB	2/9/2021 17:00	SMPL
M1308-18	MC0A14	937	0.169 -	ISM02.4	PPB	2/9/2021 17:02	SMPL
M1308-19	MC0A15	1427	0.2657 -	ISM02.4	PPB	2/9/2021 17:04	SMPL
M1308-20	MC0A24	1312	0.243 -	ISM02.4	PPB	2/9/2021 17:07	SMPL
M1308-21	MC0A25	5637	1.0965 -	ISM02.4	PPB	2/9/2021 17:09	SMPL
M1308-22	MC0A31	715	0.1252 -	ISM02.4	PPB	2/9/2021 17:11	SMPL
CCV049	CCV049	25778	5.0709 -	ISM02.4	PPB	2/9/2021 17:13	SMPL
CCB049	CCB049	-22	-0.0202 -	ISM02.4	PPB	2/9/2021 17:16	SMPL