


LB113082

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.3796e-004

C= -4.2869e-002

Rho= 0.9999548

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.030	-0.030	55	0.000	55					-15
0.2	0.200	0.170	-0.030	896	0.0 %	896					2
2.5	2.500	2.553	0.053	10909	0.0 %	10909					1
5.0	5.000	5.034	0.034	21333	0.0 %	21333					0
7.5	7.500	7.512	0.012	31747	0.0 %	31747					0
10.0	10.000	9.961	-0.039	42039	0.0 %	42039					

LB113082 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Feb 2021 11:37:16

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	55	-	0.0000	7470A	PPB	10 Feb 2021 11:38:47	S	Std
0.2 - 1	S0.2	896	-	0.2000	7470A	PPB	10 Feb 2021 11:41:05	S	Std
2.5 - 1	S2.5	10909	-	2.5000	7470A	PPB	10 Feb 2021 11:43:22	S	Std
5.0 - 1	S5	21333	-	5.0000	7470A	PPB	10 Feb 2021 11:45:40	S	Std
7.5 - 1	S7.5	31747	-	7.5000	7470A	PPB	10 Feb 2021 11:47:56	S	Std
10.0 - 1	S10	42039	-	10.0000	7470A	PPB	10 Feb 2021 11:50:12	S	Std
ICV97 - 1	ICV97	16015	3.7681	-	7470A	PPB	10 Feb 2021 11:52:56	U	SMPL
ICB97 - 1	ICB97	-35	-0.0512	-	7470A	PPB	10 Feb 2021 11:55:10	U	SMPL
CCV50 - 1	CCV50	20856	4.9201	-	7470A	PPB	10 Feb 2021 11:57:25	U	SMPL
CCB50 - 1	CCB50	-21	-0.0479	-	7470A	PPB	10 Feb 2021 11:59:40	U	SMPL
CRA - 1	CRA	917	0.1753	-	7470A	PPB	10 Feb 2021 12:01:55	U	SMPL
HighStd - 1	HighStd	41444	9.8193	-	7470A	PPB	10 Feb 2021 12:04:11	U	SMPL
ChkStd - 1	ChkStd	28939	6.8436	-	7470A	PPB	10 Feb 2021 12:06:27	U	SMPL
PB134561BL - 1	PBW	43	-0.0326	-	7470A	PPB	10 Feb 2021 12:08:43	U	SMPL
PB134561BS - 1	LCSW	15839	3.7263	-	7470A	PPB	10 Feb 2021 12:10:59	U	SMPL
M1341-03 - 1	26920	55	-0.0298	-	7470A	PPB	10 Feb 2021 12:13:16	U	SMPL
M1341-03DUP - 1	26920DUP	103	-0.0184	-	7470A	PPB	10 Feb 2021 12:15:32	U	SMPL
M1341-03MS - 1	26920MS	17704	4.1701	-	7470A	PPB	10 Feb 2021 12:22:45	U	SMPL
M1341-03MSD - 1	26920MSD	16656	3.9207	-	7470A	PPB	10 Feb 2021 12:25:02	U	SMPL
M1341-04 - 1	26878	1018	0.1994	-	7470A	PPB	10 Feb 2021 12:27:17	U	SMPL
CCV51 - 1	CCV51	20614	4.8625	-	7470A	PPB	10 Feb 2021 12:29:33	U	SMPL
CCB51 - 1	CCB51	26	-0.0367	-	7470A	PPB	10 Feb 2021 12:31:49	U	SMPL
PB134561TB - 1	PB134538TB	103	-0.0184	-	7470A	PPB	10 Feb 2021 12:34:04	U	SMPL
M1341-03LX5 - 1		110	-0.0167	-	7470A	PPB	10 Feb 2021 12:36:20	U	SMPL
M1341-03A - 1		17292	4.0720	-	7470A	PPB	10 Feb 2021 12:38:36	U	SMPL
CCV52 - 1	CCV52	20524	4.8411	-	7470A	PPB	10 Feb 2021 12:40:52	U	SMPL
CCB52 - 1	CCB52	35	-0.0345	-	7470A	PPB	10 Feb 2021 12:43:08	U	SMPL