

LB113084


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

47453



A= 0.0000e+000

B= 2.1068e-004 slope

C= -3.9401e-002 y-intercept

Rho= 0.9999384

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.00	0.000	-0.001	-0.001	184	0.000	184					-
0.05	0.050					334					-
0.20	0.200	0.181	-0.019	1047	0.0 %	1047					-10
2.50	2.500	2.468	-0.032	11902	0.0 %	11902					-1
5.00	5.000	5.083	0.083	24312	0.0 %	24312					2
7.50	7.500	7.510	0.010	35835	0.0 %	35835					0
10.0	10.000	9.958	-0.042	47453	0.0 %	47453					0

LB113084

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	184	-		0 SFAM01.1	PPB	2/10/2021 13:33	Std
	0.2 S01	1047	-		0.2 SFAM01.1	PPB	2/10/2021 13:35	Std
	2.5 S02	11902	-		2.5 SFAM01.1	PPB	2/10/2021 13:38	Std
	5 S03	24312	-		5 SFAM01.1	PPB	2/10/2021 13:40	Std
	7.5 S04	35835	-		7.5 SFAM01.1	PPB	2/10/2021 13:42	Std
	10 S05	47453	-		10 SFAM01.1	PPB	2/10/2021 13:44	Std
ICV011	ICV011	18190	3.7929 -		SFAM01.1	PPB	2/10/2021 13:47	SMPL
ICB011	ICB98011	109	-0.0164 -		SFAM01.1	PPB	2/10/2021 13:49	SMPL
CCV053	CCV053	24173	5.0534 -		SFAM01.1	PPB	2/10/2021 13:52	SMPL
CCB053	CCB053	89	-0.0207 -		SFAM01.1	PPB	2/10/2021 13:54	SMPL
PB134562BL	PBW562	87	-0.0211 -		SFAM01.1	PPB	2/10/2021 13:56	SMPL
M1325-07	X3019	41433	8.6898 -		SFAM01.1	PPB	2/10/2021 13:58	SMPL
M1325-08	X3019D	41414	8.6858 -		SFAM01.1	PPB	2/10/2021 14:05	SMPL
M1325-09	X3019S	46258	9.7064 -		SFAM01.1	PPB	2/10/2021 14:07	SMPL
CCV054	CCV054	23826	4.9803 -		SFAM01.1	PPB	2/10/2021 14:12	SMPL
CCB054	CCB054	74	-0.0238 -		SFAM01.1	PPB	2/10/2021 14:14	SMPL